

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041919\
 Data File : VN055159.D
 Acq On : 20 Apr 2019 00:11
 Operator : JC/SP
 Sample : K2472-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EXCAVATION-NEWARK-SOUTH-PIL

Quant Time: Apr 20 05:22:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	531313	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	843135	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	686226	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	205712	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	348581	49.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.72%	
35) Dibromofluoromethane	7.59	113	275463	48.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.78%	
50) Toluene-d8	10.09	98	1008699	50.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.04%	
62) 4-Bromofluorobenzene	12.40	95	277671	42.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.70%	
Target Compounds						
16) Acetone	3.83	43	14267	7.443	ug/l	96
20) Methylene Chloride	4.55	84	130653	19.389	ug/l	96
43) Isopropyl Acetate	8.18	43	18322	1.838	ug/l #	94
84) 1,2,4-Trimethylbenzene	13.04	105	17711	1.269	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041919\
Data File : VN055159.D
Acq On : 20 Apr 2019 00:11
Operator : JC/SP
Sample : K2472-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EXCAVATION-NEWARK-SOUTH-PIL

Quant Time: Apr 20 05:22:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 12:28:28 2019
Response via : Initial Calibration

