

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041919\
 Data File : VN055161.D
 Acq On : 20 Apr 2019 1:02
 Operator : JC/SP
 Sample : K2199-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-03-041819-A

Quant Time: Apr 20 05:30:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	565476	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	881775	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	695379	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	213811	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	352756	46.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.86%	
35) Dibromofluoromethane	7.59	113	281945	47.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.68%	
50) Toluene-d8	10.09	98	1032977	48.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.96%	
62) 4-Bromofluorobenzene	12.40	95	283532	41.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.70%	
Target Compounds						
16) Acetone	3.82	43	17473	8.565	ug/l	96
20) Methylene Chloride	4.55	84	232342	32.397	ug/l	98
43) Isopropyl Acetate	8.18	43	20337	1.950	ug/l #	95
95) Naphthalene	15.13	128	15089	4.926	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041919\
Data File : VN055161.D
Acq On : 20 Apr 2019 1:02
Operator : JC/SP
Sample : K2199-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-03-041819-A

Quant Time: Apr 20 05:30:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 12:28:28 2019
Response via : Initial Calibration

