

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041921\
 Data File : VN066651.D
 Acq On : 19 Apr 2021 10:15
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Apr 20 04:52:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 15 19:16:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.580	168	262225	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.514	114	400379	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	359151	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.287	152	179263	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.945	65	148171	37.46	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	74.92%		
35) Dibromofluoromethane	7.505	113	118277	42.60	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	85.20%		
50) Toluene-d8	10.026	98	449998	41.35	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	82.70%		
62) 4-Bromofluorobenzene	12.346	95	155488	40.63	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	81.26%		
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.827	85	148893	46.75	ug/l	100
3) Chloromethane	2.029	50	205993	44.35	ug/l	99
4) Vinyl Chloride	2.160	62	190487	47.09	ug/l	98
5) Bromomethane	2.474	94	34258	13.61	ug/l	100
6) Chloroethane	2.576	64	25482	10.51	ug/l	98
7) Trichlorofluoromethane	2.879	101	220800	47.13	ug/l	99
8) Diethyl Ether	3.340	74	97228	50.77	ug/l	83
9) 1,1,2-Trichlorotrifluo...	3.640	101	140793	50.88	ug/l	99
10) Methyl Iodide	3.842	142	183953	47.86	ug/l	97
11) Tert butyl alcohol	4.815	59	126689	199.54	ug/l	99
12) 1,1-Dichloroethene	3.624	96	138256	51.02	ug/l	90
13) Acrolein	3.530	56	60808	179.72	ug/l	97
14) Allyl chloride	4.212	41	288569	247.09	ug/l	93
15) Acrylonitrile	4.893	53	417063	238.65	ug/l	99
16) Acetone	3.764	43	448808	319.76	ug/l	97
17) Carbon Disulfide	3.933	76	413963	45.32	ug/l	99
18) Methyl Acetate	4.241	43	222362	52.19	ug/l	95
19) Methyl tert-butyl Ether	4.947	73	498353	51.93	ug/l	96
20) Methylene Chloride	4.440	84	173027	46.29	ug/l	95
21) trans-1,2-Dichloroethene	4.920	96	155508	50.97	ug/l	91
22) Diisopropyl ether	5.858	45	597549	51.57	ug/l	97
23) Vinyl Acetate	5.791	43	2436182	251.73	ug/l	96
24) 1,1-Dichloroethane	5.735	63	309697	49.08	ug/l	98
25) 2-Butanone	6.754	43	543436	233.05	ug/l	93
26) 2,2-Dichloropropane	6.719	77	259262	51.83	ug/l	98
27) cis-1,2-Dichloroethene	6.725	96	180445	51.05	ug/l	91
28) Bromochloromethane	7.100	49	143514	43.86	ug/l	85
29) Tetrahydrofuran	7.127	42	365096	232.65	ug/l	92
30) Chloroform	7.283	83	298203	50.24	ug/l	97
31) Cyclohexane	7.561	56	273803	48.13	ug/l	90
32) 1,1,1-Trichloroethane	7.481	97	245578	50.09	ug/l	97
36) 1,1-Dichloropropene	7.704	75	219689	54.44	ug/l	99
37) Ethyl Acetate	6.754	43	543436	118.75	ug/l #	68
38) Carbon Tetrachloride	7.685	117	205928	54.47	ug/l	98
39) Methylcyclohexane	9.007	83	260458	56.68	ug/l	94
40) Benzene	7.958	78	704578	54.51	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041921\
 Data File : VN066651.D
 Acq On : 19 Apr 2021 10:15
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Apr 20 04:52:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 15 19:16:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.124	41	326623	155.98	ug/l #	100
42) 1,2-Dichloroethane	8.044	62	230962	52.04	ug/l	100
43) Isopropyl Acetate	8.095	43	378968	49.58	ug/l #	88
44) Trichloroethene	8.763	130	164585	55.13	ug/l	98
45) 1,2-Dichloropropane	9.047	63	196154	54.51	ug/l	99
46) Dibromomethane	9.138	93	112406	52.08	ug/l	100
47) Bromodichloromethane	9.337	83	238032	53.72	ug/l	100
48) Methyl methacrylate	9.138	41	174605	50.79	ug/l	90
49) 1,4-Dioxane	9.147	88	52785	905.13	ug/l	90
51) 4-Methyl-2-Pentanone	9.927	43	1063116	236.38	ug/l	96
52) Toluene	10.091	92	420440	55.21	ug/l	99
53) t-1,3-Dichloropropene	10.321	75	258084	53.94	ug/l	98
54) cis-1,3-Dichloropropene	9.774	75	293702	55.71	ug/l	95
55) 1,1,2-Trichloroethane	10.504	97	163611	52.00	ug/l	98
56) Ethyl methacrylate	10.375	69	229456	49.50	ug/l	94
57) 1,3-Dichloropropane	10.646	76	278523	52.81	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.632	63	279186	369.71	ug/l	94
59) 2-Hexanone	10.699	43	719193	224.92	ug/l	97
60) Dibromochloromethane	10.844	129	176240	53.69	ug/l	99
61) 1,2-Dibromoethane	10.946	107	161771	52.33	ug/l	100
64) Tetrachloroethene	10.568	164	140827	56.34	ug/l	99
65) Chlorobenzene	11.375	112	420430	53.81	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.450	131	163621	57.72	ug/l	99
67) Ethyl Benzene	11.453	91	772421	56.80	ug/l	98
68) m/p-Xylenes	11.566	106	577967	116.45	ug/l	99
69) o-Xylene	11.890	106	278322	57.47	ug/l	99
70) Styrene	11.909	104	456495	60.49	ug/l	99
71) Bromoform	12.073	173	123903	57.92	ug/l #	98
73) Isopropylbenzene	12.193	105	737356	54.01	ug/l	100
74) N-amyl acetate	12.022	43	203415	49.71	ug/l	91
75) 1,1,2,2-Tetrachloroethane	12.451	83	242391	47.71	ug/l	99
76) 1,2,3-Trichloropropane	12.499	75	283698	67.67	ug/l #	100
77) Bromobenzene	12.469	156	183775	54.56	ug/l	94
78) n-propylbenzene	12.534	91	847725	53.71	ug/l	98
79) 2-Chlorotoluene	12.617	91	502690	52.08	ug/l	99
80) 1,3,5-Trimethylbenzene	12.679	105	620427	53.86	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.247	75	69871	52.17	ug/l	98
82) 4-Chlorotoluene	12.716	91	519268	54.67	ug/l	97
83) tert-Butylbenzene	12.936	119	513056	52.93	ug/l	99
84) 1,2,4-Trimethylbenzene	12.984	105	601348	55.30	ug/l	98
85) sec-Butylbenzene	13.116	105	701126	53.53	ug/l	100
86) p-Isopropyltoluene	13.234	119	608193	56.58	ug/l	100
87) 1,3-Dichlorobenzene	13.226	146	320002	53.74	ug/l	99
88) 1,4-Dichlorobenzene	13.306	146	329140	55.48	ug/l	99
89) n-Butylbenzene	13.558	91	501611	55.75	ug/l	99
90) Hexachloroethane	13.816	117	124219	52.25	ug/l	93
91) 1,2-Dichlorobenzene	13.596	146	324861	55.02	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.215	75	40907	47.42	ug/l	95
93) 1,2,4-Trichlorobenzene	14.851	180	185730	55.87	ug/l	98
94) Hexachlorobutadiene	14.948	225	102861	54.75	ug/l	99
95) Naphthalene	15.071	128	435651	47.61	ug/l	100
96) 1,2,3-Trichlorobenzene	15.248	180	175053	51.24	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041921\
Data File : VN066651.D
Acq On : 19 Apr 2021 10:15
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :


Quant Time: Apr 20 04:52:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 15 19:16:35 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 20 04:52:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 15 19:16:35 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.580 min Scan# 2184
Delta R.T. -0.008 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:168 Resp: 262225
Ion Ratio Lower Upper
168 100
99 52.5 42.6 63.8



#2
Dichlorodifluoromethane
Concen: 46.75 ug/l
RT: 1.827 min Scan# 39
Delta R.T. -0.006 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion: 85 Resp: 148893
Ion Ratio Lower Upper
85 100
87 33.0 16.6 49.8









#3
Chloromethane
Concen: 44.35 ug/l
RT: 2.029 min Scan# 114
Delta R.T. -0.002 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 50 Resp: 205993
Ion Ratio Lower Upper
50 100
52 32.9 25.8 38.8



#4
Vinyl Chloride
Concen: 47.09 ug/l
RT: 2.160 min Scan# 163
Delta R.T. -0.000 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion: 62 Resp: 190487
Ion Ratio Lower Upper
62 100
64 32.3 25.1 37.7









#5
Bromomethane
Concen: 13.61 ug/l
RT: 2.474 min Scan# 280
Delta R.T. -0.013 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 94 Resp: 34258
Ion Ratio Lower Upper
94 100
96 91.5 73.0 109.4



#6
Chloroethane
Concen: 10.51 ug/l
RT: 2.576 min Scan# 318
Delta R.T. -0.067 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion: 64 Resp: 25482
Ion Ratio Lower Upper
64 100
66 33.0 25.4 38.2









#7
Trichlorofluoromethane
Concen: 47.13 ug/l
RT: 2.879 min Scan# 431
Delta R.T. -0.083 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:101 Resp: 220800
Ion Ratio Lower Upper
101 100
103 64.2 52.1 78.1



#8
Diethyl Ether
Concen: 50.77 ug/l
RT: 3.340 min Scan# 603
Delta R.T. -0.014 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion: 74 Resp: 97228
Ion Ratio Lower Upper
74 100
45 109.2 46.4 139.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.88 ug/l
RT: 3.640 min Scan# 715
Delta R.T. -0.046 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:101 Resp: 140793
Ion Ratio Lower Upper
101 100
85 43.9 34.6 51.8
151 79.2 63.5 95.3



#10
Methyl Iodide
Concen: 47.86 ug/l
RT: 3.842 min Scan# 790
Delta R.T. -0.037 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion:142 Resp: 183953
Ion Ratio Lower Upper
142 100
127 37.5 31.7 47.5
141 13.7 10.9 16.3









#11

Tert butyl alcohol
Concen: 199.54 ug/l
RT: 4.815 min Scan# 1153
Delta R.T. 0.099 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Instrument :
MSVOA_N
ClientSampled :




Tgt Ion: 59 Resp: 126689
Ion Ratio Lower Upper
59 100
57 9.6 8.0 12.0




#12

1,1-Dichloroethene
Concen: 51.02 ug/l
RT: 3.624 min Scan# 709
Delta R.T. -0.043 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion: 96 Resp: 138256
Ion Ratio Lower Upper
96 100
61 170.9 122.5 183.7
98 64.9 50.8 76.2









#13
Acrolein
Concen: 179.72 ug/l
RT: 3.530 min Scan# 674
Delta R.T. -0.014 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 56 Resp: 60808
Ion Ratio Lower Upper
56 100
55 72.7 56.0 84.0



#14
Allyl chloride
Concen: 247.09 ug/l
RT: 4.212 min Scan# 928
Delta R.T. -0.032 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15


Tgt Ion: 41 Resp: 288569
Ion Ratio Lower Upper
41 100
39 62.1 53.0 79.4
76 30.7 28.9 43.3








#15
Acrylonitrile
Concen: 238.65 ug/l
RT: 4.893 min Scan# 1182
Delta R.T. -0.000 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15


Instrument :
MSVOA_N
ClientSampled :





Tgt Ion: 53 Resp: 417063
Ion Ratio Lower Upper
53 100
52 82.0 66.1 99.1
51 34.7 27.9 41.9




#16
Acetone
Concen: 319.76 ug/l
RT: 3.764 min Scan# 761
Delta R.T. 0.011 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15



Tgt Ion: 43 Resp: 448808
Ion Ratio Lower Upper
43 100
58 32.1 27.0 40.6









#17
Carbon Disulfide
Concen: 45.32 ug/l
RT: 3.933 min Scan# 824
Delta R.T. -0.045 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 76 Resp: 413963
Ion Ratio Lower Upper
76 100
78 9.2 7.8 11.6



#18
Methyl Acetate
Concen: 52.19 ug/l
RT: 4.241 min Scan# 939
Delta R.T. -0.008 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion: 43 Resp: 222362
Ion Ratio Lower Upper
43 100
74 21.8 19.5 29.3









#19

Methyl tert-butyl Ether
 Concen: 51.93 ug/l
 RT: 4.947 min Scan# 1202
 Delta R.T. -0.008 min
 Lab File: VN066651.D
 Acq: 19 Apr 2021 10:15

Instrument :
 MSVOA_N
 ClientSampleId :




Tgt Ion: 73 Resp: 498353
 Ion Ratio Lower Upper
 73 100
 57 23.5 17.3 25.9




#20

Methylene Chloride
 Concen: 46.29 ug/l
 RT: 4.440 min Scan# 1013
 Delta R.T. -0.027 min
 Lab File: VN066651.D
 Acq: 19 Apr 2021 10:15


Tgt Ion: 84 Resp: 173027
 Ion Ratio Lower Upper
 84 100
 49 126.3 94.6 141.8
 51 38.8 29.9 44.9
 86 61.6 51.4 77.2








#21
trans-1,2-Dichloroethene
Concen: 50.97 ug/l
RT: 4.920 min Scan# 1192
Delta R.T. -0.029 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15


Instrument :
MSVOA_N
ClientSampled :





Tgt Ion: 96 Resp: 155508
Ion Ratio Lower Upper
96 100
61 146.7 104.9 157.3
98 63.5 51.8 77.8




#22
Diisopropyl ether
Concen: 51.57 ug/l
RT: 5.858 min Scan# 1542
Delta R.T. -0.006 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15



Tgt Ion: 45 Resp: 597549
Ion Ratio Lower Upper
45 100
43 61.4 48.9 73.3
87 24.7 23.7 35.5
59 11.1 9.9 14.9









#23
 Vinyl Acetate
 Concen: 251.73 ug/l
 RT: 5.791 min Scan# 1517
 Delta R.T. -0.014 min
 Lab File: VN066651.D
 Acq: 19 Apr 2021 10:15




Instrument :
 MSVOA_N
 ClientSampled :


Tgt Ion: 43 Resp: 2436182
 Ion Ratio Lower Upper
 43 100
 86 9.3 8.6 13.0



#24
 1,1-Dichloroethane
 Concen: 49.08 ug/l
 RT: 5.735 min Scan# 1496
 Delta R.T. -0.022 min
 Lab File: VN066651.D
 Acq: 19 Apr 2021 10:15

Tgt Ion: 63 Resp: 309697
 Ion Ratio Lower Upper
 63 100
 98 6.1 3.5 10.4
 100 3.8 2.3 6.9









#25
2-Butanone
Concen: 233.05 ug/l
RT: 6.754 min Scan# 1876
Delta R.T. 0.008 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 543436
Ion Ratio Lower Upper
43 100
72 25.2 23.1 34.7



#26
2,2-Dichloropropane
Concen: 51.83 ug/l
RT: 6.719 min Scan# 1863
Delta R.T. -0.014 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15


Tgt Ion: 77 Resp: 259262
Ion Ratio Lower Upper
77 100
97 23.1 11.9 35.9








#27
 cis-1,2-Dichloroethene
 Concen: 51.05 ug/l
 RT: 6.725 min Scan# 1865
 Delta R.T. -0.016 min
 Lab File: VN066651.D
 Acq: 19 Apr 2021 10:15

Instrument :
 MSVOA_N
 ClientSampled :







Tgt Ion:	96	Resp:	180445
Ion	Ratio	Lower	Upper
96	100		
61	153.3	0.0	276.8
98	64.3	0.0	129.4



#28
 Bromochloromethane
 Concen: 43.86 ug/l
 RT: 7.100 min Scan# 2005
 Delta R.T. -0.011 min
 Lab File: VN066651.D
 Acq: 19 Apr 2021 10:15

Tgt Ion:	49	Resp:	143514
Ion	Ratio	Lower	Upper
49	100		
129	1.8	0.0	3.6
130	70.5	67.8	101.6





#29

Tetrahydrofuran

Concen: 232.65 ug/l

RT: 7.127 min Scan# 2015





Delta R.T. -0.003 min

Lab File: VN066651.D

Acq: 19 Apr 2021 10:15

Tgt Ion: 42 Resp: 365096

Ion	Ratio	Lower	Upper
42	100		
72	40.2	36.6	55.0
71	38.2	34.3	51.5



#30

Chloroform

Concen: 50.24 ug/l

RT: 7.283 min Scan# 2073




Delta R.T. -0.008 min


Lab File: VN066651.D

Acq: 19 Apr 2021 10:15

Tgt Ion: 83 Resp: 298203


Ion	Ratio	Lower	Upper
83	100		
85	63.5	52.6	78.8








#31
Cyclohexane
Concen: 48.13 ug/l
RT: 7.561 min Scan# 2177
Delta R.T. -0.014 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion: 56 Resp: 273803
Ion Ratio Lower Upper
56 100
69 30.0 26.6 39.8
84 77.6 70.4 105.6




#32
1,1,1-Trichloroethane
Concen: 50.09 ug/l
RT: 7.481 min Scan# 2147
Delta R.T. -0.011 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15



Tgt Ion: 97 Resp: 245578
Ion Ratio Lower Upper
97 100
99 64.9 50.8 76.2
61 46.5 34.3 51.5









#33
1,2-Dichloroethane-d4
Concen: 37.46 ug/l
RT: 7.945 min Scan# 2320
Delta R.T. -0.008 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 65 Resp: 148171
Ion Ratio Lower Upper
65 100
67 54.9 0.0 107.2



#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.514 min Scan# 2532
Delta R.T. -0.005 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion: 114 Resp: 400379
Ion Ratio Lower Upper
114 100
63 21.7 0.0 38.6
88 15.8 0.0 30.6









#35
Dibromofluoromethane
Concen: 42.60 ug/l
RT: 7.505 min Scan# 2156
Delta R.T. -0.006 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:113 Resp: 118277
Ion Ratio Lower Upper
113 100
111 101.3 82.2 123.2
192 20.5 15.4 23.0



#36
1,1-Dichloropropene
Concen: 54.44 ug/l
RT: 7.704 min Scan# 2230
Delta R.T. -0.013 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion: 75 Resp: 219689
Ion Ratio Lower Upper
75 100
110 35.6 18.2 54.6
77 31.5 24.9 37.3





#37

Ethyl Acetate

Concen: 118.75 ug/l

RT: 6.754 min Scan# 1876





Delta R.T. -0.089 min

Lab File: VN066651.D

Acq: 19 Apr 2021 10:15

Tgt Ion:	43	Resp:	543436
Ion	Ratio	Lower	Upper
43	100		
61	0.0	11.6	17.4#
70	0.0	8.6	13.0#

Instrument :
MSVOA_N
ClientSampled :



#38

Carbon Tetrachloride

Concen: 54.47 ug/l




RT: 7.685 min Scan# 2223


Delta R.T. -0.011 min

Lab File: VN066651.D

Acq: 19 Apr 2021 10:15

Tgt Ion:	117	Resp:	205928
Ion	Ratio	Lower	Upper
117	100		
119	97.1	76.0	114.0
121	30.6	24.5	36.7









#39
Methylcyclohexane
Concen: 56.68 ug/l
RT: 9.007 min Scan# 2716
Delta R.T. -0.008 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 83 Resp: 260458
Ion Ratio Lower Upper
83 100
55 81.1 58.9 88.3
98 44.2 35.9 53.9



#40
Benzene
Concen: 54.51 ug/l
RT: 7.958 min Scan# 2325
Delta R.T. -0.011 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15


Tgt Ion: 78 Resp: 704578
Ion Ratio Lower Upper
78 100
77 23.6 19.0 28.4








#41
Methacrylonitrile
Concen: 155.98 ug/l
RT: 7.124 min Scan# 2014
Delta R.T. 0.037 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Instrument :
MSVOA_N
ClientSampled :







Tgt Ion: 41 Resp: 326623
Ion Ratio Lower Upper
41 100
39 0.0 0.0 0.0
67 0.0 0.0 0.0
52 0.0 0.0 0.0



#42
1,2-Dichloroethane
Concen: 52.04 ug/l
RT: 8.044 min Scan# 2357
Delta R.T. -0.008 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15


Tgt Ion: 62 Resp: 230962
Ion Ratio Lower Upper
62 100
98 9.1 0.0 18.6








#43
Isopropyl Acetate
Concen: 49.58 ug/l
RT: 8.095 min Scan# 2376
Delta R.T. 0.000 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Instrument :
MSVOA_N
ClientSampled :







Tgt Ion: 43 Resp: 378968
Ion Ratio Lower Upper
43 100
61 20.1 22.5 33.7#
87 12.2 11.1 16.7



#44
Trichloroethene
Concen: 55.13 ug/l
RT: 8.763 min Scan# 2625
Delta R.T. -0.005 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion:130 Resp: 164585
Ion Ratio Lower Upper
130 100
95 101.5 0.0 199.6









#45
1,2-Dichloropropane
Concen: 54.51 ug/l
RT: 9.047 min Scan# 2731
Delta R.T. -0.006 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 63 Resp: 196154
Ion Ratio Lower Upper
63 100
65 30.8 24.9 37.3



#46
Dibromomethane
Concen: 52.08 ug/l
RT: 9.138 min Scan# 2765
Delta R.T. -0.003 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion: 93 Resp: 112406
Ion Ratio Lower Upper
93 100
95 83.7 67.0 100.6
174 103.9 83.2 124.8









#47
Bromodichloromethane
Concen: 53.72 ug/l
RT: 9.337 min Scan# 2839
Delta R.T. -0.000 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 238032
Ion Ratio Lower Upper
83 100
85 63.8 51.0 76.4
127 8.1 7.0 10.6



#48
Methyl methacrylate
Concen: 50.79 ug/l
RT: 9.138 min Scan# 2765
Delta R.T. 0.002 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion: 41 Resp: 174605
Ion Ratio Lower Upper
41 100
69 82.6 75.9 113.9
39 54.9 46.6 69.8









#49
1,4-Dioxane
Concen: 905.13 ug/l
RT: 9.147 min Scan# 2768
Delta R.T. 0.011 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 88 Resp: 52785
Ion Ratio Lower Upper
88 100
43 35.3 24.7 37.1
58 74.6 52.6 78.8



#50
Toluene-d8
Concen: 41.35 ug/l
RT: 10.026 min Scan# 3096
Delta R.T. -0.003 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion: 98 Resp: 449998
Ion Ratio Lower Upper
98 100
100 64.6 50.5 75.7









#51
4-Methyl-2-Pentanone
Concen: 236.38 ug/l
RT: 9.927 min Scan# 3059
Delta R.T. 0.005 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 1063116
Ion Ratio Lower Upper
43 100
58 38.3 32.5 48.7



#52
Toluene
Concen: 55.21 ug/l
RT: 10.091 min Scan# 3120
Delta R.T. -0.002 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion: 92 Resp: 420440
Ion Ratio Lower Upper
92 100
91 175.0 140.7 211.1









#53
t-1,3-Dichloropropene
Concen: 53.94 ug/l
RT: 10.321 min Scan# 3206
Delta R.T. 0.000 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 75 Resp: 258084
Ion Ratio Lower Upper
75 100
77 31.5 26.2 39.4



#54
cis-1,3-Dichloropropene
Concen: 55.71 ug/l
RT: 9.774 min Scan# 3002
Delta R.T. -0.003 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion: 75 Resp: 293702
Ion Ratio Lower Upper
75 100
77 31.3 24.9 37.3
39 50.2 36.2 54.4





#55

1,1,2-Trichloroethane

Concen: 52.00 ug/l

RT: 10.504 min Scan# 3274

Delta R.T. 0.003 min

Lab File: VN066651.D

Acq: 19 Apr 2021 10:15

Tgt Ion: 97 Resp: 163611





Ion Ratio Lower Upper

97 100

83 87.3 67.7 101.5

85 57.4 44.5 66.7

99 61.9 49.4 74.0



#56

Ethyl methacrylate

Concen: 49.50 ug/l

RT: 10.375 min Scan# 3226

Delta R.T. 0.003 min

Lab File: VN066651.D

Acq: 19 Apr 2021 10:15




Tgt Ion: 69 Resp: 229456


Ion Ratio Lower Upper

69 100

41 68.8 49.6 74.4

39 35.4 28.0 42.0









#57
1,3-Dichloropropane
Concen: 52.81 ug/l
RT: 10.646 min Scan# 3327
Delta R.T. -0.000 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 76 Resp: 278523
Ion Ratio Lower Upper
76 100
78 32.5 25.7 38.5



#58
2-Chloroethyl Vinyl ether
Concen: 369.71 ug/l
RT: 9.632 min Scan# 2949
Delta R.T. -0.000 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion: 63 Resp: 279186
Ion Ratio Lower Upper
63 100
106 24.8 22.6 33.8









#59
2-Hexanone
Concen: 224.92 ug/l
RT: 10.699 min Scan# 3347
Delta R.T. 0.005 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 719193
Ion Ratio Lower Upper
43 100
58 53.9 28.1 84.3



#60
Dibromochloromethane
Concen: 53.69 ug/l
RT: 10.844 min Scan# 3401
Delta R.T. 0.002 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion: 129 Resp: 176240
Ion Ratio Lower Upper
129 100
127 77.4 38.5 115.5









#61
1,2-Dibromoethane
Concen: 52.33 ug/l
RT: 10.946 min Scan# 3439
Delta R.T. 0.000 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:107 Resp: 161771
Ion Ratio Lower Upper
107 100
109 94.2 75.6 113.4



#62
4-Bromofluorobenzene
Concen: 40.63 ug/l
RT: 12.346 min Scan# 3961
Delta R.T. 0.002 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion: 95 Resp: 155488
Ion Ratio Lower Upper
95 100
174 82.3 0.0 169.0
176 81.2 0.0 162.4









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.348 min Scan# 3589
Delta R.T. -0.001 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:117 Resp: 359151
Ion Ratio Lower Upper
117 100
82 56.5 43.7 65.5
119 32.2 26.1 39.1



#64
Tetrachloroethene
Concen: 56.34 ug/l
RT: 10.568 min Scan# 3298
Delta R.T. -0.003 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion:164 Resp: 140827
Ion Ratio Lower Upper
164 100
166 128.0 102.6 154.0
129 94.0 76.2 114.4
131 91.6 73.4 110.2









#65
Chlorobenzene
Concen: 53.81 ug/l
RT: 11.375 min Scan# 3599
Delta R.T. 0.000 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:112 Resp: 420430
Ion Ratio Lower Upper
112 100
114 32.0 26.1 39.1



#66
1,1,1,2-Tetrachloroethane
Concen: 57.72 ug/l
RT: 11.450 min Scan# 3627
Delta R.T. 0.000 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion:131 Resp: 163621
Ion Ratio Lower Upper
131 100
133 94.4 48.0 144.0
119 68.5 34.1 102.2









#67
Ethyl Benzene
Concen: 56.80 ug/l
RT: 11.453 min Scan# 3628
Delta R.T. 0.000 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 91 Resp: 772421
Ion Ratio Lower Upper
91 100
106 30.8 25.5 38.3



#68
m/p-Xylenes
Concen: 116.45 ug/l
RT: 11.566 min Scan# 3670
Delta R.T. 0.003 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion:106 Resp: 577967
Ion Ratio Lower Upper
106 100
91 203.7 161.4 242.0









#69
o-Xylene
Concen: 57.47 ug/l
RT: 11.890 min Scan# 3791
Delta R.T. 0.000 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:106 Resp: 278322
Ion Ratio Lower Upper
106 100
91 214.7 106.8 320.4



#70
Styrene
Concen: 60.49 ug/l
RT: 11.909 min Scan# 3798
Delta R.T. -0.000 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion:104 Resp: 456495
Ion Ratio Lower Upper
104 100
78 50.3 39.0 58.4
103 55.2 44.5 66.7









#71
Bromoform
Concen: 57.92 ug/l
RT: 12.073 min Scan# 3859
Delta R.T. 0.003 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:173 Resp: 123903
Ion Ratio Lower Upper
173 100
175 48.5 23.8 71.2
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.287 min Scan# 4312
Delta R.T. -0.001 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion:152 Resp: 179263
Ion Ratio Lower Upper
152 100
115 57.4 28.5 85.6
150 170.6 0.0 347.6









#73
Isopropylbenzene
Concen: 54.01 ug/l
RT: 12.193 min Scan# 3904
Delta R.T. 0.000 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:105 Resp: 737356
Ion Ratio Lower Upper
105 100
120 26.4 13.3 39.9



#74
N-ethyl acetate
Concen: 49.71 ug/l
RT: 12.022 min Scan# 3840
Delta R.T. 0.003 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion: 43 Resp: 203415
Ion Ratio Lower Upper
43 100
70 35.1 34.4 51.6
55 23.0 20.7 31.1
61 20.5 19.1 28.7









#75
 1,1,2,2-Tetrachloroethane
 Concen: 47.71 ug/l
 RT: 12.451 min Scan# 4000
 Delta R.T. 0.003 min
 Lab File: VN066651.D
 Acq: 19 Apr 2021 10:15




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 83 Resp: 242391
 Ion Ratio Lower Upper
 83 100
 131 9.4 5.3 15.8
 85 63.6 32.0 96.0



#76
 1,2,3-Trichloropropane
 Concen: 67.67 ug/l
 RT: 12.499 min Scan# 4018
 Delta R.T. 0.003 min
 Lab File: VN066651.D
 Acq: 19 Apr 2021 10:15

Tgt Ion: 75 Resp: 283698
 Ion Ratio Lower Upper
 75 100
 77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 54.56 ug/l

RT: 12.469 min Scan# 4007





Delta R.T. -0.001 min

Lab File: VN066651.D

Acq: 19 Apr 2021 10:15

Tgt Ion:	156	Resp:	183775
Ion Ratio	Lower	Upper	
156	100		
77	205.3	96.3	288.8
158	97.1	49.0	147.0

Instrument :
MSVOA_N
ClientSampleId :



#78

n-propylbenzene

Concen: 53.71 ug/l




RT: 12.534 min Scan# 4031


Delta R.T. -0.000 min

Lab File: VN066651.D

Acq: 19 Apr 2021 10:15

Tgt Ion:	91	Resp:	847725
Ion Ratio	Lower	Upper	
91	100		
120	22.8	11.8	35.3









#79
2-Chlorotoluene
Concen: 52.08 ug/l
RT: 12.617 min Scan# 4062
Delta R.T. -0.000 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 91 Resp: 502690
Ion Ratio Lower Upper
91 100
126 34.2 17.4 52.3



#80
1,3,5-Trimethylbenzene
Concen: 53.86 ug/l
RT: 12.679 min Scan# 4085
Delta R.T. 0.003 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion:105 Resp: 620427
Ion Ratio Lower Upper
105 100
120 48.5 24.3 72.8









#81
trans-1,4-Dichloro-2-butene
Concen: 52.17 ug/l
RT: 12.247 min Scan# 3924
Delta R.T. 0.003 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 75 Resp: 69871
Ion Ratio Lower Upper
75 100
53 90.5 74.6 111.8
89 45.3 35.8 53.8



#82
4-Chlorotoluene
Concen: 54.67 ug/l
RT: 12.716 min Scan# 4099
Delta R.T. 0.000 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion: 91 Resp: 519268
Ion Ratio Lower Upper
91 100
126 33.0 17.2 51.6









#83
tert-Butylbenzene
Concen: 52.93 ug/l
RT: 12.936 min Scan# 4181
Delta R.T. 0.000 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:119 Resp: 513056
Ion Ratio Lower Upper
119 100
91 66.9 33.0 98.9
134 24.7 12.7 38.1



#84
1,2,4-Trimethylbenzene
Concen: 55.30 ug/l
RT: 12.984 min Scan# 4199
Delta R.T. 0.002 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion:105 Resp: 601348
Ion Ratio Lower Upper
105 100
120 46.1 22.4 67.0









#85
 sec-Butylbenzene
 Concen: 53.53 ug/l
 RT: 13.116 min Scan# 4248
 Delta R.T. -0.000 min
 Lab File: VN066651.D
 Acq: 19 Apr 2021 10:15




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:105 Resp: 701126
 Ion Ratio Lower Upper
 105 100
 134 19.8 10.0 30.0



#86
 p-Isopropyltoluene
 Concen: 56.58 ug/l
 RT: 13.234 min Scan# 4292
 Delta R.T. -0.000 min
 Lab File: VN066651.D
 Acq: 19 Apr 2021 10:15

Tgt Ion:119 Resp: 608193
 Ion Ratio Lower Upper
 119 100
 134 26.4 13.3 39.8
 91 23.4 11.8 35.3









#87
1,3-Dichlorobenzene
Concen: 53.74 ug/l
RT: 13.226 min Scan# 4289
Delta R.T. -0.000 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:146 Resp: 320002
Ion Ratio Lower Upper
146 100
111 39.5 19.5 58.5
148 64.0 32.3 96.9



#88
1,4-Dichlorobenzene
Concen: 55.48 ug/l
RT: 13.306 min Scan# 4319
Delta R.T. 0.000 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion:146 Resp: 329140
Ion Ratio Lower Upper
146 100
111 37.3 19.0 57.0
148 64.2 31.8 95.3









#89
n-Butylbenzene
Concen: 55.75 ug/l
RT: 13.558 min Scan# 4413
Delta R.T. 0.000 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 91 Resp: 501611
Ion Ratio Lower Upper
91 100
92 51.3 26.1 78.3
134 25.0 13.0 38.9



#90
Hexachloroethane
Concen: 52.25 ug/l
RT: 13.816 min Scan# 4509
Delta R.T. 0.003 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion: 117 Resp: 124219
Ion Ratio Lower Upper
117 100
201 81.6 37.8 113.3









#91
1,2-Dichlorobenzene
Concen: 55.02 ug/l
RT: 13.596 min Scan# 4427
Delta R.T. -0.000 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:146 Resp: 324861
Ion Ratio Lower Upper
146 100
111 40.6 20.1 60.3
148 63.7 32.0 96.0



#92
1,2-Dibromo-3-Chloropropane
Concen: 47.42 ug/l
RT: 14.215 min Scan# 4658
Delta R.T. -0.001 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion: 75 Resp: 40907
Ion Ratio Lower Upper
75 100
155 82.2 44.1 132.4
157 108.0 56.0 168.2









#93
1,2,4-Trichlorobenzene
Concen: 55.87 ug/l
RT: 14.851 min Scan# 4895
Delta R.T. 0.000 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:180 Resp: 185730
Ion Ratio Lower Upper
180 100
182 96.0 47.3 142.0
145 28.8 15.6 46.8



#94
Hexachlorobutadiene
Concen: 54.75 ug/l
RT: 14.948 min Scan# 4931
Delta R.T. -0.000 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion:225 Resp: 102861
Ion Ratio Lower Upper
225 100
223 62.9 31.9 95.9
227 63.5 31.6 95.0









#95
Naphthalene
Concen: 47.61 ug/l
RT: 15.071 min Scan# 4977
Delta R.T. -0.000 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Instrument :
MSVOA_N
ClientSampled :

Tgt Ion:128 Resp: 435651
Ion Ratio Lower Upper
128 100
127 12.6 10.2 15.2
129 10.7 8.6 13.0



#96
1,2,3-Trichlorobenzene
Concen: 51.24 ug/l
RT: 15.248 min Scan# 5043
Delta R.T. -0.003 min
Lab File: VN066651.D
Acq: 19 Apr 2021 10:15

Tgt Ion:180 Resp: 175053
Ion Ratio Lower Upper
180 100
182 96.3 47.4 142.3
145 30.8 16.5 49.5

