

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041921\
 Data File : VN066661.D
 Acq On : 19 Apr 2021 14:42
 Operator : JC/MD
 Sample : M2075-14
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-4-S

Quant Time: Apr 20 04:56:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 15 19:16:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	289488	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	459348	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	437042	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	188608	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	211125	48.34	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.68%
35) Dibromofluoromethane	7.513	113	168826	53.00	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.00%
50) Toluene-d8	10.029	98	667808	53.49	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.98%
62) 4-Bromofluorobenzene	12.346	95	215567	49.10	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.20%
Target Compounds						
					Qvalue	
16) Acetone	3.758	43	38329	16.62	ug/l	96
18) Methyl Acetate	4.257	43	5282	1.12	ug/l #	80
20) Methylene Chloride	4.474	84	9992	2.42	ug/l	97
43) Isopropyl Acetate	8.095	43	25286	2.88	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041921\
Data File : VN066661.D
Acq On : 19 Apr 2021 14:42
Operator : JC/MD
Sample : M2075-14
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-4-S

Quant Time: Apr 20 04:56:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 15 19:16:35 2021
Response via : Initial Calibration

