

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041921\
 Data File : VN066662.D
 Acq On : 19 Apr 2021 15:06
 Operator : JC/MD
 Sample : M2075-17
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-4-D

Quant Time: Apr 20 04:56:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 15 19:16:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	289719	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	455555	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	434614	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.287	152	185779	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	209475	47.93	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.86%
35) Dibromofluoromethane	7.513	113	168109	53.21	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.42%
50) Toluene-d8	10.029	98	661889	53.46	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.92%
62) 4-Bromofluorobenzene	12.346	95	214985	49.38	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.76%
Target Compounds						
					Qvalue	
16) Acetone	3.756	43	36039	15.08	ug/l	97
18) Methyl Acetate	4.263	43	6048	1.28	ug/l #	54
20) Methylene Chloride	4.477	84	8953	2.17	ug/l	92
43) Isopropyl Acetate	8.095	43	26816	3.08	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041921\
Data File : VN066662.D
Acq On : 19 Apr 2021 15:06
Operator : JC/MD
Sample : M2075-17
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-4-D

Quant Time: Apr 20 04:56:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 15 19:16:35 2021
Response via : Initial Calibration

