

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041921\
 Data File : VN066665.D
 Acq On : 19 Apr 2021 16:17
 Operator : JC/MD
 Sample : M2075-26
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-8-S

Manual Integrations
 APPROVED

MMDadoda
 4/21/2021 4:31:06 PM

Quant Time: Apr 20 04:57:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 15 19:16:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	287586	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	452912	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	429954	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	176774	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	206972	47.71	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.42%		
35) Dibromofluoromethane	7.513	113	163348	52.01	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	104.02%		
50) Toluene-d8	10.029	98	649434	52.76	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	105.52%		
62) 4-Bromofluorobenzene	12.346	95	209453	48.39	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	96.78%		
Target Compounds						Qvalue
16) Acetone	3.758	43	35158	14.67	ug/l	99
18) Methyl Acetate	4.252	43	6403m	1.37	ug/l	
20) Methylene Chloride	4.466	84	8942	2.18	ug/l	98
43) Isopropyl Acetate	8.095	43	26637	3.08	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041921\
Data File : VN066665.D
Acq On : 19 Apr 2021 16:17
Operator : JC/MD
Sample : M2075-26
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-8-S

Manual Integrations
APPROVED

MMDadoda
4/21/2021 4:31:06 PM

Quant Time: Apr 20 04:57:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 15 19:16:35 2021
Response via : Initial Calibration

