

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041921\
 Data File : VN066666.D
 Acq On : 19 Apr 2021 16:40
 Operator : JC/MD
 Sample : M2075-29
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-8-D

Quant Time: Apr 20 04:57:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 15 19:16:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.594	168	292902	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	461478	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	440957	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.287	152	183125	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	214334	48.51	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.02%
35) Dibromofluoromethane	7.513	113	172181	53.80	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	107.60%
50) Toluene-d8	10.029	98	679201	54.15	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.30%
62) 4-Bromofluorobenzene	12.346	95	216659	49.12	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.24%
Target Compounds						
					Qvalue	
16) Acetone	3.758	43	37870	16.02	ug/l	95
18) Methyl Acetate	4.260	43	6429	1.35	ug/l #	85
20) Methylene Chloride	4.469	84	9362	2.24	ug/l	95
43) Isopropyl Acetate	8.095	43	26599	3.02	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041921\
Data File : VN066666.D
Acq On : 19 Apr 2021 16:40
Operator : JC/MD
Sample : M2075-29
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-8-D

Quant Time: Apr 20 04:57:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 15 19:16:35 2021
Response via : Initial Calibration

