

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041921\
 Data File : VN066654.D
 Acq On : 19 Apr 2021 11:59
 Operator : JC/MD
 Sample : VN0419WBSD01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0419WBSD01

Manual Integrations
 APPROVED

MMDadoda
 4/21/2021 4:30:58 PM

Quant Time: Apr 20 04:53:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 15 19:16:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.588	168	239283	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.516	114	382899	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	349023	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.287	152	154417	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.950	65	166332	46.08	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	92.16%		
35) Dibromofluoromethane	7.510	113	130289	49.07	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	98.14%		
50) Toluene-d8	10.026	98	498167	47.87	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	95.74%		
62) 4-Bromofluorobenzene	12.346	95	166790	45.58	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	91.16%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.835	85	47236	16.25	ug/l	95
3) Chloromethane	2.034	50	66903	15.78	ug/l	100
4) Vinyl Chloride	2.163	62	62592	16.96	ug/l	99
5) Bromomethane	2.503	94	38304	16.67	ug/l	99
6) Chloroethane	2.653	64	36524	16.50	ug/l	98
7) Trichlorofluoromethane	2.967	101	68171	15.95	ug/l	98
8) Diethyl Ether	3.356	74	30692	17.56	ug/l	84
9) 1,1,2-Trichlorotrifluo...	3.691	101	42793	16.95	ug/l #	88
10) Methyl Iodide	3.884	142	62569	17.84	ug/l	99
11) Tert butyl alcohol	4.702	59	42283	72.98	ug/l #	91
12) 1,1-Dichloroethene	3.675	96	41807	16.91	ug/l	95
13) Acrolein	3.546	56	25423	82.34	ug/l	99
14) Allyl chloride	4.246	41	87876	82.46	ug/l	93
15) Acrylonitrile	4.890	53	128067	80.31	ug/l	99
16) Acetone	3.753	43	104408	74.96	ug/l	97
17) Carbon Disulfide	3.981	76	127435	15.29	ug/l	97
18) Methyl Acetate	4.249	43	66224	17.03	ug/l	95
19) Methyl tert-butyl Ether	4.954	73	152548	17.42	ug/l	92
20) Methylene Chloride	4.466	84	57264	16.79	ug/l	94
21) trans-1,2-Dichloroethene	4.952	96	47913	17.21	ug/l	92
22) Diisopropyl ether	5.861	45	184747	17.47	ug/l	93
23) Vinyl Acetate	5.802	43	736856	83.44	ug/l	96
24) 1,1-Dichloroethane	5.756	63	97671	16.96	ug/l	99
25) 2-Butanone	6.746	43	156567	73.58	ug/l	95
26) 2,2-Dichloropropane	6.733	77	78136	17.12	ug/l	98
27) cis-1,2-Dichloroethene	6.735	96	54961	17.04	ug/l	91
28) Bromochloromethane	7.108	49	47969	16.07	ug/l	83
29) Tetrahydrofuran	7.124	42	109540	76.50	ug/l	92
30) Chloroform	7.290	83	92761	17.13	ug/l	100
31) Cyclohexane	7.575	56	82908	15.97	ug/l	90
32) 1,1,1-Trichloroethane	7.489	97	73527	16.44	ug/l	96
36) 1,1-Dichloropropene	7.714	75	66052	17.12	ug/l	98
37) Ethyl Acetate	6.840	43	72479	16.56	ug/l	95
38) Carbon Tetrachloride	7.695	117	62930	17.41	ug/l	99
39) Methylcyclohexane	9.012	83	74017	16.84	ug/l	91
40) Benzene	7.966	78	215924	17.47	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041921\
 Data File : VN066654.D
 Acq On : 19 Apr 2021 11:59
 Operator : JC/MD
 Sample : VN0419WBSD01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0419WBSD01

Manual Integrations
 APPROVED

MMDadoda
 4/21/2021 4:30:58 PM

Quant Time: Apr 20 04:53:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 15 19:16:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.100	41	36321m	18.14	ug/l	
42) 1,2-Dichloroethane	8.049	62	72996	17.20	ug/l	100
43) Isopropyl Acetate	8.092	43	117430	16.07	ug/l #	88
44) Trichloroethene	8.768	130	50493	17.68	ug/l	98
45) 1,2-Dichloropropane	9.050	63	60283	17.52	ug/l	99
46) Dibromomethane	9.141	93	35632	17.26	ug/l	97
47) Bromodichloromethane	9.337	83	75153	17.74	ug/l	99
48) Methyl methacrylate	9.133	41	52148	15.86	ug/l	90
49) 1,4-Dioxane	9.133	88	16190	279.83	ug/l #	88
51) 4-Methyl-2-Pentanone	9.919	43	332589	77.33	ug/l	96
52) Toluene	10.090	92	128799	17.68	ug/l	100
53) t-1,3-Dichloropropene	10.321	75	78434	17.14	ug/l	96
54) cis-1,3-Dichloropropene	9.777	75	88434	17.54	ug/l	96
55) 1,1,2-Trichloroethane	10.501	97	53600	17.81	ug/l	98
56) Ethyl methacrylate	10.369	69	68289	16.02	ug/l	94
57) 1,3-Dichloropropane	10.646	76	86777	17.21	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.632	63	46368	68.56	ug/l	94
59) 2-Hexanone	10.694	43	196523	67.74	ug/l	96
60) Dibromochloromethane	10.841	129	55822	17.78	ug/l	100
61) 1,2-Dibromoethane	10.946	107	52174	17.65	ug/l	98
64) Tetrachloroethene	10.571	164	43142	17.76	ug/l	100
65) Chlorobenzene	11.375	112	134350	17.69	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.450	131	51463	18.68	ug/l	98
67) Ethyl Benzene	11.453	91	232179	17.57	ug/l	98
68) m/p-Xylenes	11.563	106	175617	36.41	ug/l	99
69) o-Xylene	11.890	106	86025	18.28	ug/l	98
70) Styrene	11.909	104	131875	17.98	ug/l	98
71) Bromoform	12.072	173	38209	18.38	ug/l #	100
73) Isopropylbenzene	12.193	105	219571	18.67	ug/l	100
74) N-amyl acetate	12.032	43	26701m	15.82	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.448	83	77139	17.62	ug/l	99
76) 1,2,3-Trichloropropane	12.499	75	66305m	18.36	ug/l	
77) Bromobenzene	12.469	156	56874	19.60	ug/l	97
78) n-propylbenzene	12.536	91	243649	17.92	ug/l	99
79) 2-Chlorotoluene	12.617	91	154254	18.55	ug/l	99
80) 1,3,5-Trimethylbenzene	12.679	105	185952	18.74	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.249	75	17134	16.36	ug/l	98
82) 4-Chlorotoluene	12.716	91	146737	17.94	ug/l	97
83) tert-Butylbenzene	12.936	119	151461	18.14	ug/l	98
84) 1,2,4-Trimethylbenzene	12.984	105	173237	18.49	ug/l	98
85) sec-Butylbenzene	13.116	105	199929	17.72	ug/l	99
86) p-Isopropyltoluene	13.234	119	169837	18.34	ug/l	99
87) 1,3-Dichlorobenzene	13.228	146	91702	17.88	ug/l	100
88) 1,4-Dichlorobenzene	13.309	146	94563	18.13	ug/l	98
89) n-Butylbenzene	13.561	91	143617	18.53	ug/l	97
90) Hexachloroethane	13.816	117	36633	17.89	ug/l	97
91) 1,2-Dichlorobenzene	13.598	146	91791	18.05	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.218	75	11731	15.79	ug/l	88
93) 1,2,4-Trichlorobenzene	14.854	180	53867	18.81	ug/l	98
94) Hexachlorobutadiene	14.947	225	28866	17.84	ug/l	99
95) Naphthalene	15.074	128	132184	16.77	ug/l	99
96) 1,2,3-Trichlorobenzene	15.251	180	51431	17.48	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041921\
Data File : VN066654.D
Acq On : 19 Apr 2021 11:59
Operator : JC/MD
Sample : VN0419WBSD01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01

Manual Integrations
APPROVED

MMDadoda
4/21/2021 4:30:58 PM

Quant Time: Apr 20 04:53:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 15 19:16:35 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041921\
Data File : VN066654.D
Acq On : 19 Apr 2021 11:59
Operator : JC/MD
Sample : VN0419WBSD01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01

Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM

Quant Time: Apr 20 04:53:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 15 19:16:35 2021
Response via : Initial Calibration







#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.588 min Scan# 2187
Delta R.T. 0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Tgt Ion: 168 Resp: 239283
Ion Ratio Lower Upper
168 100
99 51.7 42.6 63.8




Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01

Manual Integrations
APPROVED



MMDadoda
4/21/2021 4:30:58 PM




#2
Dichlorodifluoromethane
Concen: 16.25 ug/l
RT: 1.835 min Scan# 42
Delta R.T. 0.002 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59



Tgt Ion: 85 Resp: 47236
Ion Ratio Lower Upper
85 100
87 30.6 16.6 49.8









#3
Chloromethane
Concen: 15.78 ug/l
RT: 2.034 min Scan# 116
Delta R.T. 0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 50 Resp: 66903
Ion Ratio Lower Upper
50 100
52 32.2 25.8 38.8

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM




#4
Vinyl Chloride
Concen: 16.96 ug/l
RT: 2.163 min Scan# 164
Delta R.T. 0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 62 Resp: 62592
Ion Ratio Lower Upper
62 100
64 32.2 25.1 37.7









#5
Bromomethane
Concen: 16.67 ug/l
RT: 2.503 min Scan# 291
Delta R.T. 0.016 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59





Tgt Ion: 94 Resp: 38304
Ion Ratio Lower Upper
94 100
96 91.8 73.0 109.4



#6
Chloroethane
Concen: 16.50 ug/l
RT: 2.653 min Scan# 347
Delta R.T. 0.010 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Tgt Ion: 64 Resp: 36524
Ion Ratio Lower Upper
64 100
66 32.6 25.4 38.2



Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01

Manual Integrations
APPROVED

MMDadoda
4/21/2021 4:30:58 PM







#7
Trichlorofluoromethane
Concen: 15.95 ug/l
RT: 2.967 min Scan# 464
Delta R.T. 0.005 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion:101 Resp: 68171
Ion Ratio Lower Upper
101 100
103 63.8 52.1 78.1

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM



#8
Diethyl Ether
Concen: 17.56 ug/l
RT: 3.356 min Scan# 609
Delta R.T. 0.002 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 74 Resp: 30692
Ion Ratio Lower Upper
74 100
45 108.3 46.4 139.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 16.95 ug/l
RT: 3.691 min Scan# 734
Delta R.T. 0.005 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion:101 Resp: 42793
Ion Ratio Lower Upper
101 100
85 22.8 34.6 51.8#
151 78.9 63.5 95.3

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM



#10
Methyl Iodide
Concen: 17.84 ug/l
RT: 3.884 min Scan# 806
Delta R.T. 0.005 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion:142 Resp: 62569
Ion Ratio Lower Upper
142 100
127 39.2 31.7 47.5
141 13.9 10.9 16.3









#11
Tert butyl alcohol
Concen: 72.98 ug/l
RT: 4.702 min Scan# 1111
Delta R.T. -0.014 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 59 Resp: 42283
Ion Ratio Lower Upper
59 100
57 6.7 8.0 12.0#

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM



#12
1,1-Dichloroethene
Concen: 16.91 ug/l
RT: 3.675 min Scan# 728
Delta R.T. 0.008 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 96 Resp: 41807
Ion Ratio Lower Upper
96 100
61 161.5 122.5 183.7
98 65.1 50.8 76.2









#13
Acrolein
Concen: 82.34 ug/l
RT: 3.546 min Scan# 680
Delta R.T. 0.002 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 56 Resp: 25423
Ion Ratio Lower Upper
56 100
55 71.2 56.0 84.0

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM



#14
Allyl chloride
Concen: 82.46 ug/l
RT: 4.246 min Scan# 941
Delta R.T. 0.002 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 41 Resp: 87876
Ion Ratio Lower Upper
41 100
39 61.8 53.0 79.4
76 31.0 28.9 43.3









#15
Acrylonitrile
Concen: 80.31 ug/l
RT: 4.890 min Scan# 1181
Delta R.T. -0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion:	53	Resp:	128067
Ion	Ratio	Lower	Upper
53	100		
52	81.9	66.1	99.1
51	35.8	27.9	41.9

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM



#16
Acetone
Concen: 74.96 ug/l
RT: 3.753 min Scan# 757
Delta R.T. -0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion:	43	Resp:	104408
Ion	Ratio	Lower	Upper
43	100		
58	31.9	27.0	40.6









#17
Carbon Disulfide
Concen: 15.29 ug/l
RT: 3.981 min Scan# 842
Delta R.T. 0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 76 Resp: 127435
Ion Ratio Lower Upper
76 100
78 8.7 7.8 11.6

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM



#18
Methyl Acetate
Concen: 17.03 ug/l
RT: 4.249 min Scan# 942
Delta R.T. 0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 43 Resp: 66224
Ion Ratio Lower Upper
43 100
74 21.7 19.5 29.3









#19
Methyl tert-butyl Ether
Concen: 17.42 ug/l
RT: 4.954 min Scan# 1205
Delta R.T. -0.001 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 73 Resp: 152548
Ion Ratio Lower Upper
73 100
57 25.2 17.3 25.9

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM



#20
Methylene Chloride
Concen: 16.79 ug/l
RT: 4.466 min Scan# 1023
Delta R.T. -0.001 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 84 Resp: 57264
Ion Ratio Lower Upper
84 100
49 129.6 94.6 141.8
51 39.4 29.9 44.9
86 64.8 51.4 77.2









#21
trans-1,2-Dichloroethene
Concen: 17.21 ug/l
RT: 4.952 min Scan# 1204
Delta R.T. 0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion:	96	Resp:	47913
Ion	Ratio	Lower	Upper
96	100		
61	143.8	104.9	157.3
98	62.9	51.8	77.8

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM



#22
Diisopropyl ether
Concen: 17.47 ug/l
RT: 5.861 min Scan# 1543
Delta R.T. -0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion:	45	Resp:	184747
Ion	Ratio	Lower	Upper
45	100		
43	66.3	48.9	73.3
87	24.9	23.7	35.5
59	11.6	9.9	14.9









#23
 Vinyl Acetate
 Concen: 83.44 ug/l
 RT: 5.802 min Scan# 1521
 Delta R.T. -0.003 min
 Lab File: VN066654.D
 Acq: 19 Apr 2021 11:59

Tgt Ion: 43 Resp: 736856
 Ion Ratio Lower Upper
 43 100
 86 9.2 8.6 13.0

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0419WBSD01




Manual Integrations
 APPROVED


MMDadoda
 4/21/2021 4:30:58 PM



#24
 1,1-Dichloroethane
 Concen: 16.96 ug/l
 RT: 5.756 min Scan# 1504
 Delta R.T. -0.001 min
 Lab File: VN066654.D
 Acq: 19 Apr 2021 11:59

Tgt Ion: 63 Resp: 97671
 Ion Ratio Lower Upper
 63 100
 98 6.5 3.5 10.4
 100 4.0 2.3 6.9









#25
2-Butanone
Concen: 73.58 ug/l
RT: 6.746 min Scan# 1873
Delta R.T. 0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 43 Resp: 156567
Ion Ratio Lower Upper
43 100
72 26.1 23.1 34.7

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM




#26
2,2-Dichloropropane
Concen: 17.12 ug/l
RT: 6.733 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 77 Resp: 78136
Ion Ratio Lower Upper
77 100
97 23.0 11.9 35.9









#27
cis-1,2-Dichloroethene
Concen: 17.04 ug/l
RT: 6.735 min Scan# 1869
Delta R.T. -0.006 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59





Tgt Ion: 96 Resp: 54961
Ion Ratio Lower Upper
96 100
61 155.1 0.0 276.8
98 64.5 0.0 129.4



#28
Bromochloromethane
Concen: 16.07 ug/l
RT: 7.108 min Scan# 2008
Delta R.T. -0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Tgt Ion: 49 Resp: 47969
Ion Ratio Lower Upper
49 100
129 0.8 0.0 3.6
130 69.3 67.8 101.6



Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01

Manual Integrations
APPROVED

MMDadoda
4/21/2021 4:30:58 PM



#29
Tetrahydrofuran
Concen: 76.50 ug/l
RT: 7.124 min Scan# 2014
Delta R.T. -0.006 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59





Tgt Ion: 42 Resp: 109540

Ion	Ratio	Lower	Upper
42	100		
72	41.0	36.6	55.0
71	37.8	34.3	51.5

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01

Manual Integrations
APPROVED




MMDadoda
4/21/2021 4:30:58 PM




#30
Chloroform
Concen: 17.13 ug/l
RT: 7.290 min Scan# 2076
Delta R.T. -0.001 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 83 Resp: 92761

Ion	Ratio	Lower	Upper
83	100		
85	65.8	52.6	78.8









#31
Cyclohexane
Concen: 15.97 ug/l
RT: 7.575 min Scan# 2182
Delta R.T. -0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 56 Resp: 82908
Ion Ratio Lower Upper
56 100
69 35.7 26.6 39.8
84 77.1 70.4 105.6

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM



#32
1,1,1-Trichloroethane
Concen: 16.44 ug/l
RT: 7.489 min Scan# 2150
Delta R.T. -0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 97 Resp: 73527
Ion Ratio Lower Upper
97 100
99 64.9 50.8 76.2
61 47.9 34.3 51.5









#33
1,2-Dichloroethane-d4
Concen: 46.08 ug/l
RT: 7.950 min Scan# 2322
Delta R.T. -0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 65 Resp: 166332
Ion Ratio Lower Upper
65 100
67 54.9 0.0 107.2

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM




#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.516 min Scan# 2533
Delta R.T. -0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 114 Resp: 382899
Ion Ratio Lower Upper
114 100
63 21.2 0.0 38.6
88 15.6 0.0 30.6









#35
Dibromofluoromethane
Concen: 49.07 ug/l
RT: 7.510 min Scan# 2158
Delta R.T. -0.001 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59





Tgt Ion:113 Resp: 130289
Ion Ratio Lower Upper
113 100
111 104.7 82.2 123.2
192 20.6 15.4 23.0



#36
1,1-Dichloropropene
Concen: 17.12 ug/l
RT: 7.714 min Scan# 2234
Delta R.T. -0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Tgt Ion: 75 Resp: 66052
Ion Ratio Lower Upper
75 100
110 34.5 18.2 54.6
77 30.6 24.9 37.3



Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01

Manual Integrations
APPROVED

MMDadoda
4/21/2021 4:30:58 PM



#37

Ethyl Acetate

Concen: 16.56 ug/l

RT: 6.840 min Scan# 1908

Delta R.T. -0.003 min

Lab File: VN066654.D





Acq: 19 Apr 2021 11:59

Tgt Ion:	43	Resp:	72479
Ion	Ratio	Lower	Upper
43	100		
61	12.3	11.6	17.4
70	9.1	8.6	13.0

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01

Manual Integrations
APPROVED

MMDadoda
4/21/2021 4:30:58 PM



#38

Carbon Tetrachloride

Concen: 17.41 ug/l




RT: 7.695 min Scan# 2227


Delta R.T. -0.001 min

Lab File: VN066654.D

Acq: 19 Apr 2021 11:59

Tgt Ion:	117	Resp:	62930
Ion	Ratio	Lower	Upper
117	100		
119	96.6	76.0	114.0
121	30.7	24.5	36.7









#39
Methylcyclohexane
Concen: 16.84 ug/l
RT: 9.012 min Scan# 2718
Delta R.T. -0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion:	83	Resp:	74017
Ion	Ratio	Lower	Upper
83	100		
55	83.3	58.9	88.3
98	42.0	35.9	53.9

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM




#40
Benzene
Concen: 17.47 ug/l
RT: 7.966 min Scan# 2328
Delta R.T. -0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion:	78	Resp:	215924
Ion	Ratio	Lower	Upper
78	100		
77	23.8	19.0	28.4









#41
Methacrylonitrile
Concen: 18.14 ug/l m
RT: 7.100 min Scan# 2005
Delta R.T. 0.013 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59





Tgt Ion: 41 Resp: 36321
Ion Ratio Lower Upper
41 100
39 47.0 0.0 0.0#
67 73.8 0.0 0.0#
52 30.7 0.0 0.0#



#42
1,2-Dichloroethane
Concen: 17.20 ug/l
RT: 8.049 min Scan# 2359
Delta R.T. -0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Tgt Ion: 62 Resp: 72996
Ion Ratio Lower Upper
62 100
98 9.2 0.0 18.6



Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01

Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM



#43

Isopropyl Acetate
Concen: 16.07 ug/l
RT: 8.092 min Scan# 2375
Delta R.T. -0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01



Tgt Ion: 43 Resp: 117430
Ion Ratio Lower Upper
43 100
61 20.0 22.5 33.7#
87 12.1 11.1 16.7


Manual Integrations
APPROVED

MMDadoda
4/21/2021 4:30:58 PM






#44

Trichloroethene
Concen: 17.68 ug/l
RT: 8.768 min Scan# 2627
Delta R.T. 0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59



Tgt Ion: 130 Resp: 50493
Ion Ratio Lower Upper
130 100
95 101.3 0.0 199.6





#45





1,2-Dichloropropane
Concen: 17.52 ug/l
RT: 9.050 min Scan# 2732
Delta R.T. -0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 63 Resp: 60283
Ion Ratio Lower Upper
63 100
65 30.6 24.9 37.3

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01

Manual Integrations
APPROVED




MMDadoda
4/21/2021 4:30:58 PM




#46

Dibromomethane
Concen: 17.26 ug/l
RT: 9.141 min Scan# 2766
Delta R.T. 0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 93 Resp: 35632
Ion Ratio Lower Upper
93 100
95 82.8 67.0 100.6
174 100.1 83.2 124.8









#47
Bromodichloromethane
Concen: 17.74 ug/l
RT: 9.337 min Scan# 2839
Delta R.T. -0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion:	83	Resp:	75153
Ion Ratio	Lower	Upper	
83	100		
85	63.1	51.0	76.4
127	7.9	7.0	10.6

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM



#48
Methyl methacrylate
Concen: 15.86 ug/l
RT: 9.133 min Scan# 2763
Delta R.T. -0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion:	41	Resp:	52148
Ion Ratio	Lower	Upper	
41	100		
69	84.8	75.9	113.9
39	51.9	46.6	69.8









#49
1,4-Dioxane
Concen: 279.83 ug/l
RT: 9.133 min Scan# 2763
Delta R.T. -0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 88 Resp: 16190
Ion Ratio Lower Upper
88 100
43 37.8 24.7 37.1#
58 75.3 52.6 78.8

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM



#50
Toluene-d8
Concen: 47.87 ug/l
RT: 10.026 min Scan# 3096
Delta R.T. -0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59


Tgt Ion: 98 Resp: 498167
Ion Ratio Lower Upper
98 100
100 64.2 50.5 75.7





#51
4-Methyl-2-Pentanone
Concen: 77.33 ug/l
RT: 9.919 min Scan# 3056
Delta R.T. -0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01



Tgt Ion: 43 Resp: 332589
Ion Ratio Lower Upper
43 100
58 38.2 32.5 48.7




Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM



#52
Toluene
Concen: 17.68 ug/l
RT: 10.090 min Scan# 3120
Delta R.T. -0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59


Tgt Ion: 92 Resp: 128799
Ion Ratio Lower Upper
92 100
91 175.3 140.7 211.1





#53
t-1,3-Dichloropropene
Concen: 17.14 ug/l
RT: 10.321 min Scan# 3206
Delta R.T. 0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01



Tgt Ion: 75 Resp: 78434
Ion Ratio Lower Upper
75 100
77 30.8 26.2 39.4




Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM




#54
cis-1,3-Dichloropropene
Concen: 17.54 ug/l
RT: 9.777 min Scan# 3003
Delta R.T. -0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 75 Resp: 88434
Ion Ratio Lower Upper
75 100
77 32.1 24.9 37.3
39 49.2 36.2 54.4






#55
1,1,2-Trichloroethane
Concen: 17.81 ug/l
RT: 10.501 min Scan# 3273
Delta R.T. -0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Tgt Ion:	97	Resp:	53600
Ion	Ratio	Lower	Upper
97	100		
83	87.2	67.7	101.5
85	57.8	44.5	66.7
99	62.1	49.4	74.0




Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01

Manual Integrations
APPROVED



MMDadoda
4/21/2021 4:30:58 PM




#56
Ethyl methacrylate
Concen: 16.02 ug/l
RT: 10.369 min Scan# 3224
Delta R.T. -0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59



Tgt Ion:	69	Resp:	68289
Ion	Ratio	Lower	Upper
69	100		
41	68.7	49.6	74.4
39	36.3	28.0	42.0









#57
1,3-Dichloropropane
Concen: 17.21 ug/l
RT: 10.646 min Scan# 3327
Delta R.T. -0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 76 Resp: 86777
Ion Ratio Lower Upper
76 100
78 32.8 25.7 38.5

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM



#58
2-Chloroethyl Vinyl ether
Concen: 68.56 ug/l
RT: 9.632 min Scan# 2949
Delta R.T. -0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59


Tgt Ion: 63 Resp: 46368
Ion Ratio Lower Upper
63 100
106 25.1 22.6 33.8





#59
2-Hexanone
Concen: 67.74 ug/l
RT: 10.694 min Scan# 3345
Delta R.T. -0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Tgt Ion: 43 Resp: 196523
Ion Ratio Lower Upper
43 100
58 53.3 28.1 84.3

Manual Integrations
APPROVED



MMDadoda
4/21/2021 4:30:58 PM




#60
Dibromochloromethane
Concen: 17.78 ug/l
RT: 10.841 min Scan# 3400
Delta R.T. -0.001 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Tgt Ion: 129 Resp: 55822
Ion Ratio Lower Upper
129 100
127 76.8 38.5 115.5





#61
1,2-Dibromoethane
Concen: 17.65 ug/l
RT: 10.946 min Scan# 3439
Delta R.T. 0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Tgt Ion:107 Resp: 52174
Ion Ratio Lower Upper
107 100
109 92.2 75.6 113.4

Manual Integrations
APPROVED



MMDadoda
4/21/2021 4:30:58 PM




#62
4-Bromofluorobenzene
Concen: 45.58 ug/l
RT: 12.346 min Scan# 3961
Delta R.T. 0.002 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Tgt Ion: 95 Resp: 166790
Ion Ratio Lower Upper
95 100
174 83.3 0.0 169.0
176 80.4 0.0 162.4





#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.348 min Scan# 3589
Delta R.T. -0.001 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Tgt Ion:117 Resp: 349023
Ion Ratio Lower Upper
117 100
82 56.7 43.7 65.5
119 32.8 26.1 39.1

Manual Integrations
APPROVED



MMDadoda
4/21/2021 4:30:58 PM




#64
Tetrachloroethene
Concen: 17.76 ug/l
RT: 10.571 min Scan# 3299
Delta R.T. -0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Tgt Ion:164 Resp: 43142
Ion Ratio Lower Upper
164 100
166 127.6 102.6 154.0
129 94.6 76.2 114.4
131 91.9 73.4 110.2







#65
Chlorobenzene
Concen: 17.69 ug/l
RT: 11.375 min Scan# 3599
Delta R.T. 0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Tgt Ion:112 Resp: 134350
Ion Ratio Lower Upper
112 100
114 31.7 26.1 39.1




Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01

Manual Integrations
APPROVED



MMDadoda
4/21/2021 4:30:58 PM




#66
1,1,1,2-Tetrachloroethane
Concen: 18.68 ug/l
RT: 11.450 min Scan# 3627
Delta R.T. 0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Tgt Ion:131 Resp: 51463
Ion Ratio Lower Upper
131 100
133 97.2 48.0 144.0
119 70.3 34.1 102.2





#67
Ethyl Benzene
Concen: 17.57 ug/l
RT: 11.453 min Scan# 3628
Delta R.T. -0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01



Tgt Ion: 91 Resp: 232179
Ion Ratio Lower Upper
91 100
106 30.9 25.5 38.3




Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM



#68
m/p-Xylenes
Concen: 36.41 ug/l
RT: 11.563 min Scan# 3669
Delta R.T. -0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 106 Resp: 175617
Ion Ratio Lower Upper
106 100
91 203.0 161.4 242.0









#69
o-Xylene
Concen: 18.28 ug/l
RT: 11.890 min Scan# 3791
Delta R.T. 0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion:106 Resp: 86025
Ion Ratio Lower Upper
106 100
91 211.2 106.8 320.4

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM



#70
Styrene
Concen: 17.98 ug/l
RT: 11.909 min Scan# 3798
Delta R.T. -0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion:104 Resp: 131875
Ion Ratio Lower Upper
104 100
78 51.1 39.0 58.4
103 55.9 44.5 66.7









#71
Bromoform
Concen: 18.38 ug/l
RT: 12.072 min Scan# 3859
Delta R.T. 0.002 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion:173 Resp: 38209
Ion Ratio Lower Upper
173 100
175 47.6 23.8 71.2
254 0.0 0.0 0.0

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.287 min Scan# 4312
Delta R.T. -0.001 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59


Tgt Ion:152 Resp: 154417
Ion Ratio Lower Upper
152 100
115 56.7 28.5 85.6
150 157.0 0.0 347.6





#73
Isopropylbenzene
Concen: 18.67 ug/l
RT: 12.193 min Scan# 3904
Delta R.T. 0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Tgt Ion: 105 Resp: 219571
Ion Ratio Lower Upper
105 100
120 26.5 13.3 39.9

Manual Integrations
APPROVED



MMDadoda
4/21/2021 4:30:58 PM




#74
N-ethyl acetate
Concen: 15.82 ug/l m
RT: 12.032 min Scan# 3844
Delta R.T. 0.013 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Tgt Ion: 43 Resp: 26701
Ion Ratio Lower Upper
43 100
70 29.1 34.4 51.6#
55 17.8 20.7 31.1#
61 14.2 19.1 28.7#





#75
1,1,2,2-Tetrachloroethane
Concen: 17.62 ug/l
RT: 12.448 min Scan# 3999
Delta R.T. -0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Tgt Ion: 83 Resp: 77139
Ion Ratio Lower Upper
83 100
131 9.0 5.3 15.8
85 63.8 32.0 96.0

Manual Integrations
APPROVED



MMDadoda
4/21/2021 4:30:58 PM




#76
1,2,3-Trichloropropane
Concen: 18.36 ug/l m
RT: 12.499 min Scan# 4018
Delta R.T. 0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Tgt Ion: 75 Resp: 66305
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0









#77
Bromobenzene
Concen: 19.60 ug/l
RT: 12.469 min Scan# 4007
Delta R.T. -0.001 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59





Tgt Ion:156 Resp: 56874
Ion Ratio Lower Upper
156 100
77 199.1 96.3 288.8
158 96.7 49.0 147.0



#78
n-propylbenzene
Concen: 17.92 ug/l
RT: 12.536 min Scan# 4032
Delta R.T. 0.002 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Tgt Ion: 91 Resp: 243649
Ion Ratio Lower Upper
91 100
120 22.9 11.8 35.3



Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01

Manual Integrations
APPROVED

MMDadoda
4/21/2021 4:30:58 PM







#79
2-Chlorotoluene
Concen: 18.55 ug/l
RT: 12.617 min Scan# 4062
Delta R.T. -0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01

Tgt Ion: 91 Resp: 154254
Ion Ratio Lower Upper
91 100
126 34.1 17.4 52.3




Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM




#80
1,3,5-Trimethylbenzene
Concen: 18.74 ug/l
RT: 12.679 min Scan# 4085
Delta R.T. 0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion: 105 Resp: 185952
Ion Ratio Lower Upper
105 100
120 48.1 24.3 72.8









#81
trans-1,4-Dichloro-2-butene
Concen: 16.36 ug/l
RT: 12.249 min Scan# 3925
Delta R.T. 0.005 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59





Tgt Ion: 75 Resp: 17134
Ion Ratio Lower Upper
75 100
53 93.0 74.6 111.8
89 40.6 35.8 53.8



#82
4-Chlorotoluene
Concen: 17.94 ug/l
RT: 12.716 min Scan# 4099
Delta R.T. 0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Tgt Ion: 91 Resp: 146737
Ion Ratio Lower Upper
91 100
126 32.7 17.2 51.6




Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01

Manual Integrations
APPROVED




MMDadoda
4/21/2021 4:30:58 PM




#83
tert-Butylbenzene
Concen: 18.14 ug/l
RT: 12.936 min Scan# 4181
Delta R.T. 0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59





Tgt Ion:119 Resp: 151461
Ion Ratio Lower Upper
119 100
91 67.7 33.0 98.9
134 24.6 12.7 38.1



#84
1,2,4-Trimethylbenzene
Concen: 18.49 ug/l
RT: 12.984 min Scan# 4199
Delta R.T. 0.002 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Tgt Ion:105 Resp: 173237
Ion Ratio Lower Upper
105 100
120 46.2 22.4 67.0



Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01


Manual Integrations
APPROVED

MMDadoda
4/21/2021 4:30:58 PM



#85
 sec-Butylbenzene
 Concen: 17.72 ug/l
 RT: 13.116 min Scan# 4248
 Delta R.T. -0.000 min
 Lab File: VN066654.D
 Acq: 19 Apr 2021 11:59




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0419WBSD01




Tgt Ion:105 Resp: 199929
 Ion Ratio Lower Upper
 105 100
 134 19.4 10.0 30.0

Manual Integrations
 APPROVED



MMDadoda
 4/21/2021 4:30:58 PM




#86
 p-Isopropyltoluene
 Concen: 18.34 ug/l
 RT: 13.234 min Scan# 4292
 Delta R.T. -0.000 min
 Lab File: VN066654.D
 Acq: 19 Apr 2021 11:59




Tgt Ion:119 Resp: 169837
 Ion Ratio Lower Upper
 119 100
 134 26.6 13.3 39.8
 91 22.6 11.8 35.3









#87
1,3-Dichlorobenzene
Concen: 17.88 ug/l
RT: 13.228 min Scan# 4290
Delta R.T. 0.002 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59





Tgt Ion:146 Resp: 91702
Ion Ratio Lower Upper
146 100
111 39.1 19.5 58.5
148 64.3 32.3 96.9



#88
1,4-Dichlorobenzene
Concen: 18.13 ug/l
RT: 13.309 min Scan# 4320
Delta R.T. 0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Tgt Ion:146 Resp: 94563
Ion Ratio Lower Upper
146 100
111 38.2 19.0 57.0
148 60.8 31.8 95.3



Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01


Manual Integrations
APPROVED

MMDadoda
4/21/2021 4:30:58 PM



#89
n-Butylbenzene
Concen: 18.53 ug/l
RT: 13.561 min Scan# 4414
Delta R.T. 0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Tgt Ion: 91 Resp: 143617
Ion Ratio Lower Upper
91 100
92 50.8 26.1 78.3
134 23.6 13.0 38.9

Manual Integrations
APPROVED



MMDadoda
4/21/2021 4:30:58 PM




#90
Hexachloroethane
Concen: 17.89 ug/l
RT: 13.816 min Scan# 4509
Delta R.T. 0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Tgt Ion: 117 Resp: 36633
Ion Ratio Lower Upper
117 100
201 78.1 37.8 113.3





#91
1,2-Dichlorobenzene
Concen: 18.05 ug/l
RT: 13.598 min Scan# 4428
Delta R.T. 0.002 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59




Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Tgt Ion:146 Resp: 91791
Ion Ratio Lower Upper
146 100
111 41.6 20.1 60.3
148 65.5 32.0 96.0

Manual Integrations
APPROVED



MMDadoda
4/21/2021 4:30:58 PM




#92
1,2-Dibromo-3-Chloropropane
Concen: 15.79 ug/l
RT: 14.218 min Scan# 4659
Delta R.T. 0.002 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59



Tgt Ion: 75 Resp: 11731
Ion Ratio Lower Upper
75 100
155 74.3 44.1 132.4
157 102.4 56.0 168.2









#93
1,2,4-Trichlorobenzene
Concen: 18.81 ug/l
RT: 14.854 min Scan# 4896
Delta R.T. 0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion	Ratio	Lower	Upper
180	100		
182	93.8	47.3	142.0
145	28.3	15.6	46.8

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01




Manual Integrations
APPROVED


MMDadoda
4/21/2021 4:30:58 PM



#94
Hexachlorobutadiene
Concen: 17.84 ug/l
RT: 14.947 min Scan# 4931
Delta R.T. -0.001 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion	Ratio	Lower	Upper
225	100		
223	63.5	31.9	95.9
227	64.5	31.6	95.0









#95
Naphthalene
Concen: 16.77 ug/l
RT: 15.074 min Scan# 4978
Delta R.T. 0.003 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion:128 Resp: 132184
Ion Ratio Lower Upper
128 100
127 12.2 10.2 15.2
129 10.2 8.6 13.0

Instrument :
MSVOA_N
ClientSampleId :
VN0419WBSD01

Manual Integrations
APPROVED

MMDadoda
4/21/2021 4:30:58 PM



#96
1,2,3-Trichlorobenzene
Concen: 17.48 ug/l
RT: 15.251 min Scan# 5044
Delta R.T. -0.000 min
Lab File: VN066654.D
Acq: 19 Apr 2021 11:59

Tgt Ion:180 Resp: 51431
Ion Ratio Lower Upper
180 100
182 94.7 47.4 142.3
145 30.8 16.5 49.5

