

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041922\
 Data File : VN072044.D
 Acq On : 19 Apr 2022 12:59
 Operator : JC\MD
 Sample : PB144164TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB144164TB

Quant Time: Apr 20 01:24:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	521412	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	885596	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1083112	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	402259	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.428	65	281450	47.282	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.560%
35) Dibromofluoromethane	8.010	113	245506	47.111	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.220%
50) Toluene-d8	10.439	98	1209738	56.747	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	113.500%
62) 4-Bromofluorobenzene	12.727	95	434731	63.054	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	126.100%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	17474	6.869	ug/l #	85
20) Methylene Chloride	5.081	84	178806	33.348	ug/l	96
43) Isopropyl Acetate	8.557	43	56369	4.255	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041922\
Data File : VN072044.D
Acq On : 19 Apr 2022 12:59
Operator : JC\MD
Sample : PB144164TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB144164TB

Quant Time: Apr 20 01:24:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

