

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041922\
 Data File : VN072047.D
 Acq On : 19 Apr 2022 14:13
 Operator : JC\MD
 Sample : N2452-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-15-1-3-GR

Quant Time: Apr 20 01:25:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	581407	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	977711	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1241220	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	465268	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	304297	45.846	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.700%
35) Dibromofluoromethane	8.016	113	282355	49.077	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.160%
50) Toluene-d8	10.439	98	1351377	57.419	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	114.840%
62) 4-Bromofluorobenzene	12.727	95	501639	65.904	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	131.800%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	68203	24.044	ug/l	91
20) Methylene Chloride	5.081	84	157687	26.374	ug/l	93
25) 2-Butanone	7.333	43	20429	4.996	ug/l	97
43) Isopropyl Acetate	8.557	43	49939	3.414	ug/l	92
95) Naphthalene	15.498	128	7009	4.109	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041922\
Data File : VN072047.D
Acq On : 19 Apr 2022 14:13
Operator : JC\MD
Sample : N2452-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-15-1-3-GR

Quant Time: Apr 20 01:25:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

