

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041922\
 Data File : VN072053.D
 Acq On : 19 Apr 2022 16:38
 Operator : JC\MD
 Sample : N2468-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-15-1-6-GR

Quant Time: Apr 20 01:26:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	593378	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1013746	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1211134	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	431570	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.428	65	316470	46.718	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.440%
35) Dibromofluoromethane	8.016	113	274640	46.039	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.080%
50) Toluene-d8	10.439	98	1375939	56.385	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	112.760%
62) 4-Bromofluorobenzene	12.727	95	482483	61.134	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	122.260%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	74022	25.569	ug/l	94
18) Methyl Acetate	4.851	43	22257	1.640	ug/l	98
20) Methylene Chloride	5.087	84	316829	51.923	ug/l	98
43) Isopropyl Acetate	8.557	43	106229	7.005	ug/l #	75
95) Naphthalene	15.504	128	6110	4.090	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041922\
Data File : VN072053.D
Acq On : 19 Apr 2022 16:38
Operator : JC\MD
Sample : N2468-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-15-1-6-GR

Quant Time: Apr 20 01:26:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

