

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041922\
 Data File : VN072054.D
 Acq On : 19 Apr 2022 17:02
 Operator : JC\MD
 Sample : N2468-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-15-1-6

Quant Time: Apr 20 01:27:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.075	168	572895	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	981046	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1119818	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	445996	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.428	65	303067	46.339	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	92.680%		
35) Dibromofluoromethane	8.010	113	261454	45.290	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	90.580%		
50) Toluene-d8	10.433	98	1300236	55.058	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	110.120%		
62) 4-Bromofluorobenzene	12.727	95	476432	62.379	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	124.760%		
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	111463	39.878	ug/l	98
18) Methyl Acetate	4.851	43	28218	2.643	ug/l	94
20) Methylene Chloride	5.075	84	418449	71.028	ug/l	98
25) 2-Butanone	7.340	43	39591	9.825	ug/l	97
31) Cyclohexane	8.081	56	19823	2.229	ug/l #	37
39) Methylcyclohexane	9.451	83	20294	2.115	ug/l	92
43) Isopropyl Acetate	8.557	43	136714	9.315	ug/l #	71
68) m/p-Xylenes	11.945	106	46278	3.173	ug/l	94
69) o-Xylene	12.274	106	56402	3.896	ug/l	97
73) Isopropylbenzene	12.574	105	123640	3.290	ug/l	100
78) n-propylbenzene	12.916	91	72212	1.827	ug/l	98
80) 1,3,5-Trimethylbenzene	13.057	105	223296	7.499	ug/l	99
84) 1,2,4-Trimethylbenzene	13.363	105	534431	18.395	ug/l	99
85) sec-Butylbenzene	13.498	105	36701	1.040	ug/l #	74
95) Naphthalene	15.504	128	15547	4.510	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041922\
Data File : VN072054.D
Acq On : 19 Apr 2022 17:02
Operator : JC\MD
Sample : N2468-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-15-1-6

Quant Time: Apr 20 01:27:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

