

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041922\
 Data File : VN072055.D
 Acq On : 19 Apr 2022 17:26
 Operator : JC\MD
 Sample : N2468-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-15-1-7-GR

Quant Time: Apr 20 01:27:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	561490	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	949700	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1138687	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	432374	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.428	65	296692	46.285	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	92.580%		
35) Dibromofluoromethane	8.016	113	258536	46.263	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	92.520%		
50) Toluene-d8	10.433	98	1280693	56.021	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	112.040%		
62) 4-Bromofluorobenzene	12.727	95	464878	62.876	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	125.760%		
Target Compounds						Qvalue
16) Acetone	4.293	43	63787	23.285	ug/l	99
18) Methyl Acetate	4.857	43	28154	2.719	ug/l	97
20) Methylene Chloride	5.075	84	336937	58.354	ug/l	97
25) 2-Butanone	7.340	43	20240	5.125	ug/l	94
43) Isopropyl Acetate	8.557	43	106412	7.490	ug/l #	74
52) Toluene	10.498	92	18927	1.215	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041922\
Data File : VN072055.D
Acq On : 19 Apr 2022 17:26
Operator : JC\MD
Sample : N2468-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-15-1-7-GR

Quant Time: Apr 20 01:27:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

