

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041922\
 Data File : VN072058.D
 Acq On : 19 Apr 2022 18:38
 Operator : JC\MD
 Sample : N2468-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-15-1-8

Quant Time: Apr 20 01:28:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	641102	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1092290	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1265517	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	484339	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.428	65	340905	46.579	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	93.160%		
35) Dibromofluoromethane	8.016	113	296264	46.093	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	92.180%		
50) Toluene-d8	10.433	98	1447076	55.036	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	110.080%		
62) 4-Bromofluorobenzene	12.727	95	523250	61.532	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	123.060%		
Target Compounds					Qvalue	
16) Acetone	4.287	43	146132	46.720	ug/l	95
18) Methyl Acetate	4.851	43	22428	1.425	ug/l	99
20) Methylene Chloride	5.087	84	371985	56.424	ug/l	95
25) 2-Butanone	7.339	43	52412	11.623	ug/l	95
31) Cyclohexane	8.086	56	21745	2.185	ug/l #	57
39) Methylcyclohexane	9.451	83	16051	1.502	ug/l	97
43) Isopropyl Acetate	8.557	43	103683	6.345	ug/l #	75
68) m/p-Xylenes	11.939	106	22254	1.350	ug/l	88
69) o-Xylene	12.274	106	30205	1.846	ug/l	96
73) Isopropylbenzene	12.574	105	97960	2.400	ug/l	99
78) n-propylbenzene	12.916	91	49598	1.155	ug/l	99
80) 1,3,5-Trimethylbenzene	13.057	105	129192	3.995	ug/l	99
84) 1,2,4-Trimethylbenzene	13.363	105	304223	9.642	ug/l	99
85) sec-Butylbenzene	13.498	105	47579	1.242	ug/l	79
95) Naphthalene	15.504	128	10300	4.234	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041922\
Data File : VN072058.D
Acq On : 19 Apr 2022 18:38
Operator : JC\MD
Sample : N2468-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-15-1-8

Quant Time: Apr 20 01:28:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

