

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041922\
 Data File : VN072061.D
 Acq On : 19 Apr 2022 19:51
 Operator : JC\MD
 Sample : N2452-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-15-1-3-GR

Quant Time: Apr 20 01:29:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	495254	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	796592	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	744706	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	259758	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	316006	55.892	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.780%
35) Dibromofluoromethane	8.016	113	265694	56.682	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	113.360%
50) Toluene-d8	10.439	98	1071584	55.883	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	111.760%
62) 4-Bromofluorobenzene	12.727	95	326239	52.605	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	105.220%
Target Compounds						
						Qvalue
16) Acetone	4.287	43	84915	35.143	ug/l	100
18) Methyl Acetate	4.857	43	19420	1.786	ug/l #	89
20) Methylene Chloride	5.093	84	175918	34.542	ug/l	97
25) 2-Butanone	7.334	43	23522	6.753	ug/l	98
43) Isopropyl Acetate	8.557	43	97626	8.192	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041922\
Data File : VN072061.D
Acq On : 19 Apr 2022 19:51
Operator : JC\MD
Sample : N2452-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-15-1-3-GR

Quant Time: Apr 20 01:29:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

