

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041922\
 Data File : VN072064.D
 Acq On : 19 Apr 2022 21:03
 Operator : JC\MD
 Sample : N2452-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-15-1-4

Quant Time: Apr 20 01:30:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	509016	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	831729	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	758538	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	288753	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	298231	51.322	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.640%	
35) Dibromofluoromethane	8.016	113	251281	51.342	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 102.680%	
50) Toluene-d8	10.439	98	988304	49.363	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 98.720%	
62) 4-Bromofluorobenzene	12.727	95	320410	49.483	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 98.960%	
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	72934	29.368	ug/l	93
18) Methyl Acetate	4.857	43	19492	1.707	ug/l	93
20) Methylene Chloride	5.087	84	299912	57.296	ug/l	95
31) Cyclohexane	8.080	56	11637	1.473	ug/l #	4
43) Isopropyl Acetate	8.557	43	115970	9.320	ug/l #	73
67) Ethyl Benzene	11.839	91	50451	1.962	ug/l	95
68) m/p-Xylenes	11.945	106	33337	3.374	ug/l	99
69) o-Xylene	12.274	106	103666	10.572	ug/l	98
73) Isopropylbenzene	12.574	105	35905	1.476	ug/l	99
78) n-propylbenzene	12.915	91	38074	1.488	ug/l	98
80) 1,3,5-Trimethylbenzene	13.051	105	201933	10.474	ug/l	100
84) 1,2,4-Trimethylbenzene	13.363	105	308971	16.426	ug/l	100
95) Naphthalene	15.498	128	52736	7.508	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041922\
Data File : VN072064.D
Acq On : 19 Apr 2022 21:03
Operator : JC\MD
Sample : N2452-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-15-1-4

Quant Time: Apr 20 01:30:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

