

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041923\
 Data File : VN077367.D
 Acq On : 19 Apr 2023 11:30
 Operator : JC\MD
 Sample : VN0419MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0419MBL01

Quant Time: Apr 20 04:24:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	521876	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.101	114	983284	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	878288	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	312964	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	330804	46.379	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.760%
35) Dibromofluoromethane	8.172	113	305659	48.558	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.120%
50) Toluene-d8	10.566	98	1198785	47.307	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.620%
62) 4-Bromofluorobenzene	12.854	95	364366	43.234	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041923\
Data File : VN077367.D
Acq On : 19 Apr 2023 11:30
Operator : JC\MD
Sample : VN0419MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0419MBL01

Quant Time: Apr 20 04:24:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

