

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041923\
 Data File : VN077373.D
 Acq On : 19 Apr 2023 14:20
 Operator : JC\MD
 Sample : PB152199ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152199ZHE#01

Quant Time: Apr 20 04:25:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	554773	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1034881	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	943430	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	335374	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	341150	44.993	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.980%
35) Dibromofluoromethane	8.172	113	317913	47.987	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.980%
50) Toluene-d8	10.571	98	1272504	47.713	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.420%
62) 4-Bromofluorobenzene	12.854	95	399318	45.018	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.040%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	39793	14.234	ug/l	93
20) Methylene Chloride	5.289	84	35221	4.461	ug/l	94
37) Ethyl Acetate	7.566	43	91601	10.449	ug/l #	94
43) Isopropyl Acetate	8.689	43	439570	27.173	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041923\
Data File : VN077373.D
Acq On : 19 Apr 2023 14:20
Operator : JC\MD
Sample : PB152199ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152199ZHE#01

Quant Time: Apr 20 04:25:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

