

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041923\
 Data File : VN077375.D
 Acq On : 19 Apr 2023 15:09
 Operator : JC\MD
 Sample : PB152199ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152199ZHE#03

Quant Time: Apr 20 04:26:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	582206	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1105329	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	997531	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	362690	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	360886	45.353	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.700%
35) Dibromofluoromethane	8.172	113	329672	46.590	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.180%
50) Toluene-d8	10.571	98	1356205	47.610	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.220%
62) 4-Bromofluorobenzene	12.854	95	427072	45.079	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.160%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	40683	13.867	ug/l	95
20) Methylene Chloride	5.284	84	36185	4.367	ug/l	94
37) Ethyl Acetate	7.572	43	94375	10.079	ug/l	96
43) Isopropyl Acetate	8.695	43	465846	26.962	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041923\
Data File : VN077375.D
Acq On : 19 Apr 2023 15:09
Operator : JC\MD
Sample : PB152199ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152199ZHE#03

Quant Time: Apr 20 04:26:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

