

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041923\
 Data File : VN077376.D
 Acq On : 19 Apr 2023 15:33
 Operator : JC\MD
 Sample : PB152199ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152199ZHE#04

Quant Time: Apr 20 04:26:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	533708	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1020084	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	934062	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	326255	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	335355	45.974	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.940%
35) Dibromofluoromethane	8.172	113	310750	47.586	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.180%
50) Toluene-d8	10.572	98	1260218	47.938	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.880%
62) 4-Bromofluorobenzene	12.854	95	386850	44.245	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.500%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	40557	15.080	ug/l #	85
20) Methylene Chloride	5.284	84	36615	4.821	ug/l	94
37) Ethyl Acetate	7.572	43	99657	11.533	ug/l	97
43) Isopropyl Acetate	8.695	43	479243	30.055	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041923\
Data File : VN077376.D
Acq On : 19 Apr 2023 15:33
Operator : JC\MD
Sample : PB152199ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152199ZHE#04

Quant Time: Apr 20 04:26:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

