

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041923\
 Data File : VN077377.D
 Acq On : 19 Apr 2023 15:57
 Operator : JC\MD
 Sample : PB152199ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152199ZHE#05

Quant Time: Apr 20 04:26:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	553507	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1051103	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	950224	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	341347	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	345756	45.705	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.400%
35) Dibromofluoromethane	8.172	113	316784	47.079	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.160%
50) Toluene-d8	10.572	98	1273935	47.029	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.060%
62) 4-Bromofluorobenzene	12.848	95	392974	43.620	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.240%
Target Compounds						
					Qvalue	
16) Acetone	4.449	43	40900	14.664	ug/l	99
20) Methylene Chloride	5.290	84	35379	4.491	ug/l	96
37) Ethyl Acetate	7.566	43	93724	10.526	ug/l	94
43) Isopropyl Acetate	8.689	43	479569	29.188	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041923\
Data File : VN077377.D
Acq On : 19 Apr 2023 15:57
Operator : JC\MD
Sample : PB152199ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152199ZHE#05

Quant Time: Apr 20 04:26:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

