

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041923\
 Data File : VN077379.D
 Acq On : 19 Apr 2023 16:46
 Operator : JC\MD
 Sample : PB152199ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152199ZHE#07

Quant Time: Apr 20 04:26:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	542226	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1011885	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	938744	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	334584	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	333553	45.009	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.020%
35) Dibromofluoromethane	8.172	113	303557	46.861	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.720%
50) Toluene-d8	10.572	98	1254735	48.116	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.240%
62) 4-Bromofluorobenzene	12.854	95	391014	45.084	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.160%
Target Compounds						
					Qvalue	
16) Acetone	4.443	43	32972	12.067	ug/l #	86
20) Methylene Chloride	5.290	84	38890	5.040	ug/l	95
37) Ethyl Acetate	7.566	43	92152	10.750	ug/l	97
43) Isopropyl Acetate	8.695	43	495535	31.329	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041923\
Data File : VN077379.D
Acq On : 19 Apr 2023 16:46
Operator : JC\MD
Sample : PB152199ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152199ZHE#07

Quant Time: Apr 20 04:26:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

