

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041923\
 Data File : VN077383.D
 Acq On : 19 Apr 2023 18:22
 Operator : JC\MD
 Sample : PB152199ZHE#11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152199ZHE#11

Quant Time: Apr 20 04:27:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	643005	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1215308	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	1134337	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	407814	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	400237	45.543	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.080%
35) Dibromofluoromethane	8.172	113	370215	47.585	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.180%
50) Toluene-d8	10.571	98	1508676	48.170	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.340%
62) 4-Bromofluorobenzene	12.848	95	477090	45.801	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.600%
Target Compounds						
					Qvalue	
16) Acetone	4.437	43	34574	10.670	ug/l	98
20) Methylene Chloride	5.284	84	39838	4.353	ug/l	98
37) Ethyl Acetate	7.572	43	88390	8.586	ug/l	97
43) Isopropyl Acetate	8.695	43	459044	24.164	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041923\
Data File : VN077383.D
Acq On : 19 Apr 2023 18:22
Operator : JC\MD
Sample : PB152199ZHE#11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152199ZHE#11

Quant Time: Apr 20 04:27:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

