

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041923\
 Data File : VN077384.D
 Acq On : 19 Apr 2023 18:47
 Operator : JC\MD
 Sample : PB152199ZHE#12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152199ZHE#12

Quant Time: Apr 20 04:27:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	612102	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1168862	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	1082512	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	395558	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	387387	46.306	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.620%
35) Dibromofluoromethane	8.172	113	348638	46.593	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.180%
50) Toluene-d8	10.571	98	1420356	47.152	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.300%
62) 4-Bromofluorobenzene	12.854	95	445131	44.431	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.860%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	32333	10.482	ug/l	98
20) Methylene Chloride	5.289	84	38530	4.423	ug/l	95
37) Ethyl Acetate	7.566	43	88092	8.897	ug/l #	94
43) Isopropyl Acetate	8.695	43	488263	26.723	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041923\
Data File : VN077384.D
Acq On : 19 Apr 2023 18:47
Operator : JC\MD
Sample : PB152199ZHE#12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152199ZHE#12

Quant Time: Apr 20 04:27:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

