

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042021\
 Data File : VN066679.D
 Acq On : 20 Apr 2021 13:58
 Operator : JC/MD
 Sample : VN0420WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0420WBL01

Quant Time: Apr 21 10:07:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042021W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Apr 20 12:52:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.111	128	53183	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.516	114	276770	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.348	117	233864	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.950	65	116542	30.41	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	101.37%
60) 4-Bromofluorobenzene	12.349	95	84969	23.18	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	77.27%
63) Toluene-d8	10.029	98	337332	31.35	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	104.50%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042021\
Data File : VN066679.D
Acq On : 20 Apr 2021 13:58
Operator : JC/MD
Sample : VN0420WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0420WBL01

Quant Time: Apr 21 10:07:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042021W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Apr 20 12:52:17 2021
Response via : Initial Calibration

