

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042021\
 Data File : VN066682.D
 Acq On : 20 Apr 2021 15:16
 Operator : JC/MD
 Sample : M2032-02 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HOLDING-TANK-260

Quant Time: Apr 21 10:08:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042021W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Apr 20 12:52:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.111	128	51706	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.519	114	271801	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.351	117	237388	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.956	65	112406	30.17	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery = 100.57%			
60) 4-Bromofluorobenzene	12.349	95	87639	23.55	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery = 78.50%			
63) Toluene-d8	10.029	98	337014	30.86	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery = 102.87%			
Target Compounds						
15) Acetone	3.761	58	2761	6.53	ug/l	70
66) Ethyl Benzene	11.459	91	5477	0.38	ug/l	93
67) m/p-Xylenes	11.568	106	7688	1.42	ug/l	89
68) o-Xylene	11.898	106	3981	0.76	ug/l	97
76) 1,3,5-Trimethylbenzene	12.681	105	8747	0.76	ug/l	89
80) 1,2,4-Trimethylbenzene	12.993	105	7882	0.71	ug/l	98
82) p-Isopropyltoluene	13.234	119	94821	8.59	ug/l	100
91) Naphthalene	15.079	128	23173	2.95	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042021\
Data File : VN066682.D
Acq On : 20 Apr 2021 15:16
Operator : JC/MD
Sample : M2032-02 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HOLDING-TANK-260

Quant Time: Apr 21 10:08:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042021W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Apr 20 12:52:17 2021
Response via : Initial Calibration

