

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042021\
 Data File : VN066686.D
 Acq On : 20 Apr 2021 17:14
 Operator : JC/MD
 Sample : M2078-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TWP-01

Quant Time: Apr 21 10:09:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042021W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Apr 20 12:52:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.116	128	46673	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.519	114	238331	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.351	117	214270	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.956	65	100356	29.84	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.47%
60) 4-Bromofluorobenzene	12.351	95	75266	22.41	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	74.70%
63) Toluene-d8	10.029	98	301201	30.55	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.83%
Target Compounds						
						Qvalue
3) Chloromethane	2.034	50	1270	0.36	ug/l	97
15) Acetone	3.756	58	1024	2.68	ug/l #	33
18) Methylene Chloride	4.472	84	3406	1.10	ug/l	88
82) p-Isopropyltoluene	13.247	119	4500	0.45	ug/l	62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042021\
Data File : VN066686.D
Acq On : 20 Apr 2021 17:14
Operator : JC/MD
Sample : M2078-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TWP-01

Quant Time: Apr 21 10:09:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042021W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Apr 20 12:52:17 2021
Response via : Initial Calibration

