

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042022\
 Data File : VN072069.D
 Acq On : 20 Apr 2022 11:05
 Operator : JC\MD
 Sample : VN0420WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0420WBL01

Quant Time: Apr 20 15:33:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	583507	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	989111	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1029416	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	398128	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	319140	47.909	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.820%
35) Dibromofluoromethane	8.016	113	286927	49.297	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.600%
50) Toluene-d8	10.439	98	1239810	52.072	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.140%
62) 4-Bromofluorobenzene	12.727	95	434131	56.378	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	112.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042022\
Data File : VN072069.D
Acq On : 20 Apr 2022 11:05
Operator : JC\MD
Sample : VN0420WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0420WBL01

Quant Time: Apr 20 15:33:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

