

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042022\
 Data File : VN072074.D
 Acq On : 20 Apr 2022 13:28
 Operator : JC\MD
 Sample : N2468-02RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-15-1-6-GRRE

Quant Time: Apr 21 03:28:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	525118	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	834303	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	751148	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	265231	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	269810	45.007	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	90.020%
35) Dibromofluoromethane	8.016	113	231931	47.242	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.480%
50) Toluene-d8	10.439	98	901465	44.887	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	89.780%
62) 4-Bromofluorobenzene	12.727	95	273092	42.045	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	84.100%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	77303	30.173	ug/l	97
20) Methylene Chloride	5.098	84	327816	60.707	ug/l	97
43) Isopropyl Acetate	8.557	43	60187	4.822	ug/l #	88
95) Naphthalene	15.498	128	5927	4.256	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042022\
Data File : VN072074.D
Acq On : 20 Apr 2022 13:28
Operator : JC\MD
Sample : N2468-02RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-15-1-6-GRRE

Quant Time: Apr 21 03:28:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

