

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042022\
 Data File : VN072075.D
 Acq On : 20 Apr 2022 13:52
 Operator : JC\MD
 Sample : N2468-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-15-1-6

Quant Time: Apr 21 03:28:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	496368	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	790871	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	710839	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	279163	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	283269	49.989	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.980%
35) Dibromofluoromethane	8.016	113	242141	52.031	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.060%
50) Toluene-d8	10.433	98	970030	50.953	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.900%
62) 4-Bromofluorobenzene	12.727	95	318018	51.651	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.300%
Target Compounds						
					Qvalue	
16) Acetone	4.281	43	121181	50.039	ug/l	93
20) Methylene Chloride	5.092	84	445122	87.204	ug/l	97
25) 2-Butanone	7.333	43	42084	12.054	ug/l	93
39) Methylcyclohexane	9.457	83	19148	2.475	ug/l	95
43) Isopropyl Acetate	8.557	43	62250	5.261	ug/l #	87
68) m/p-Xylenes	11.945	106	45037	4.865	ug/l	97
69) o-Xylene	12.274	106	54530	5.934	ug/l	100
73) Isopropylbenzene	12.574	105	117458	4.993	ug/l	100
78) n-propylbenzene	12.915	91	68578	2.771	ug/l	99
80) 1,3,5-Trimethylbenzene	13.057	105	216616	11.622	ug/l	99
84) 1,2,4-Trimethylbenzene	13.363	105	528599	29.067	ug/l	99
85) sec-Butylbenzene	13.498	105	35605	1.612	ug/l #	75
95) Naphthalene	15.504	128	13349	4.773	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042022\
Data File : VN072075.D
Acq On : 20 Apr 2022 13:52
Operator : JC\MD
Sample : N2468-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-15-1-6

Quant Time: Apr 21 03:28:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

