

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042022\
 Data File : VN072076.D
 Acq On : 20 Apr 2022 14:16
 Operator : JC\MD
 Sample : N2468-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-15-1-7-GR

Quant Time: Apr 21 03:28:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	490001	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	790423	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	708726	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	247722	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	292418	52.274	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.540%	
35) Dibromofluoromethane	8.016	113	244878	52.648	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.300%	
50) Toluene-d8	10.439	98	956972	50.296	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 100.600%	
62) 4-Bromofluorobenzene	12.727	95	286142	46.500	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 93.000%	
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	67250	28.130	ug/l	98
20) Methylene Chloride	5.098	84	359623	71.370	ug/l	98
25) 2-Butanone	7.334	43	18479	5.362	ug/l	93
43) Isopropyl Acetate	8.563	43	65008	5.498	ug/l #	85
52) Toluene	10.498	92	19428	1.498	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042022\
Data File : VN072076.D
Acq On : 20 Apr 2022 14:16
Operator : JC\MD
Sample : N2468-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-15-1-7-GR

Quant Time: Apr 21 03:28:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

