

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042022\
 Data File : VN072078.D
 Acq On : 20 Apr 2022 15:04
 Operator : JC\MD
 Sample : N2480-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-16-1A-1-GR

Quant Time: Apr 21 03:29:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	568230	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	964227	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1038170	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	418088	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	306635	47.269	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.540%
35) Dibromofluoromethane	8.016	113	270689	47.708	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.420%
50) Toluene-d8	10.439	98	1240412	53.441	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.880%
62) 4-Bromofluorobenzene	12.727	95	449170	59.836	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	119.680%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	33772	12.182	ug/l	91
20) Methylene Chloride	5.092	84	19236	3.292	ug/l	98
39) Methylcyclohexane	9.457	83	20752	2.200	ug/l	90
43) Isopropyl Acetate	8.557	43	71465	4.954	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042022\
Data File : VN072078.D
Acq On : 20 Apr 2022 15:04
Operator : JC\MD
Sample : N2480-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-16-1A-1-GR

Quant Time: Apr 21 03:29:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

