

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042022\
 Data File : VN072079.D
 Acq On : 20 Apr 2022 15:28
 Operator : JC\MD
 Sample : N2480-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-16-1A-1

Quant Time: Apr 21 03:30:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	631991	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1084159	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1196361	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	506693	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	339806	47.098	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.200%
35) Dibromofluoromethane	8.016	113	303026	47.499	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.000%
50) Toluene-d8	10.439	98	1384282	53.042	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.080%
62) 4-Bromofluorobenzene	12.727	95	522157	61.864	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	123.720%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	54447	17.658	ug/l	90
20) Methylene Chloride	5.093	84	44362	6.826	ug/l	90
39) Methylcyclohexane	9.457	83	28233	2.662	ug/l	98
43) Isopropyl Acetate	8.557	43	66489	4.099	ug/l #	82
68) m/p-Xylenes	11.945	106	56962	3.656	ug/l	99
69) o-Xylene	12.274	106	32416	2.096	ug/l	98
80) 1,3,5-Trimethylbenzene	13.057	105	36348	1.074	ug/l	99
84) 1,2,4-Trimethylbenzene	13.363	105	104743	3.173	ug/l	98
95) Naphthalene	15.498	128	24494	4.784	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042022\
Data File : VN072079.D
Acq On : 20 Apr 2022 15:28
Operator : JC\MD
Sample : N2480-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-16-1A-1

Quant Time: Apr 21 03:30:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

