

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042022\
 Data File : VN072081.D
 Acq On : 20 Apr 2022 16:16
 Operator : JC\MD
 Sample : N2480-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-16-1A-2

Quant Time: Apr 21 03:31:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	701926	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1200163	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1289510	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	537157	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	382525	47.736	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.480%
35) Dibromofluoromethane	8.016	113	337700	47.817	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.640%
50) Toluene-d8	10.439	98	1527585	52.876	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.760%
62) 4-Bromofluorobenzene	12.727	95	578537	61.919	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	123.840%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	77269	22.563	ug/l	99
20) Methylene Chloride	5.092	84	54890	7.604	ug/l	90
25) 2-Butanone	7.328	43	33476	6.781	ug/l	100
39) Methylcyclohexane	9.457	83	38025	3.239	ug/l	96
43) Isopropyl Acetate	8.557	43	73307	4.083	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042022\
Data File : VN072081.D
Acq On : 20 Apr 2022 16:16
Operator : JC\MD
Sample : N2480-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-16-1A-2

Quant Time: Apr 21 03:31:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

