

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042022\
 Data File : VN072082.D
 Acq On : 20 Apr 2022 16:40
 Operator : JC\MD
 Sample : N2480-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-16-1A-3-GR

Quant Time: Apr 21 03:31:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	617960	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1067306	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1159810	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	503344	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	338215	47.942	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.880%
35) Dibromofluoromethane	8.022	113	301070	47.937	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.880%
50) Toluene-d8	10.439	98	1369532	53.306	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.620%
62) 4-Bromofluorobenzene	12.727	95	524197	63.086	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	126.180%
Target Compounds						
						Qvalue
16) Acetone	4.287	43	37983	12.598	ug/l	99
20) Methylene Chloride	5.087	84	43615	6.863	ug/l #	97
39) Methylcyclohexane	9.457	83	15716	1.505	ug/l	98
43) Isopropyl Acetate	8.557	43	73819	4.623	ug/l #	79
73) Isopropylbenzene	12.574	105	230790	5.441	ug/l	100
85) sec-Butylbenzene	13.498	105	87354	2.194	ug/l	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042022\
 Data File : VN072082.D
 Acq On : 20 Apr 2022 16:40
 Operator : JC\MD
 Sample : N2480-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-16-1A-3-GR

Quant Time: Apr 21 03:31:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

