

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042022\
 Data File : VN072084.D
 Acq On : 20 Apr 2022 17:27
 Operator : JC\MD
 Sample : N2480-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-16-1A-4-GR

Quant Time: Apr 21 03:32:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	687852	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1186248	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1271694	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	575585	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	379778	48.363	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.720%		
35) Dibromofluoromethane	8.016	113	330628	47.365	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.740%		
50) Toluene-d8	10.439	98	1498123	52.464	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	104.920%		
62) 4-Bromofluorobenzene	12.727	95	569195	61.633	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	123.260%		
Target Compounds					Qvalue	
16) Acetone	4.293	43	38325	11.420	ug/l	95
20) Methylene Chloride	5.099	84	37898	5.358	ug/l	95
39) Methylcyclohexane	9.457	83	13216	1.139	ug/l	95
43) Isopropyl Acetate	8.557	43	62052	3.497	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042022\
Data File : VN072084.D
Acq On : 20 Apr 2022 17:27
Operator : JC\MD
Sample : N2480-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-16-1A-4-GR

Quant Time: Apr 21 03:32:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

