

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042022\
 Data File : VN072085.D
 Acq On : 20 Apr 2022 17:52
 Operator : JC\MD
 Sample : N2480-16
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-16-1A-4

Quant Time: Apr 21 03:33:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	652361	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1102596	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1202769	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	539934	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	352853	47.379	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.760%
35) Dibromofluoromethane	8.016	113	311385	47.993	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.980%
50) Toluene-d8	10.439	98	1423762	53.643	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.280%
62) 4-Bromofluorobenzene	12.727	95	567546	66.117	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	132.240%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	59815	18.793	ug/l	89
20) Methylene Chloride	5.098	84	109897	16.382	ug/l	92
39) Methylcyclohexane	9.451	83	19812	1.837	ug/l	94
43) Isopropyl Acetate	8.557	43	74994	4.546	ug/l #	80
73) Isopropylbenzene	12.574	105	164264	3.610	ug/l	100
78) n-propylbenzene	12.916	91	102592	2.144	ug/l	100
85) sec-Butylbenzene	13.492	105	89386	2.092	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042022\
Data File : VN072085.D
Acq On : 20 Apr 2022 17:52
Operator : JC\MD
Sample : N2480-16
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-16-1A-4

Quant Time: Apr 21 03:33:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

