

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042022\
 Data File : VN072107.D
 Acq On : 21 Apr 2022 03:54
 Operator : JC\MD
 Sample : N2482-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-23R

Quant Time: Apr 21 07:14:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	634058	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1087697	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1143079	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	448281	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	338201	46.722	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.440%
35) Dibromofluoromethane	8.016	113	304736	47.611	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.220%
50) Toluene-d8	10.439	98	1365704	52.160	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.320%
62) 4-Bromofluorobenzene	12.727	95	477813	56.426	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	112.860%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	14983	4.843	ug/l	93
19) Methyl tert-butyl Ether	5.616	73	13637	0.701	ug/l #	45
37) Ethyl Acetate	7.416	43	46672	4.771	ug/l	97
40) Benzene	8.457	78	91490	3.240	ug/l	96
52) Toluene	10.498	92	9727	0.545	ug/l	98
67) Ethyl Benzene	11.839	91	49394	1.275	ug/l	99
68) m/p-Xylenes	11.951	106	23689	1.591	ug/l	99
69) o-Xylene	12.280	106	5458	0.369	ug/l	92
73) Isopropylbenzene	12.574	105	20824	0.551	ug/l	100
78) n-propylbenzene	12.916	91	22960	0.578	ug/l	96
84) 1,2,4-Trimethylbenzene	13.363	105	24682	0.845	ug/l	99
95) Naphthalene	15.504	128	36437	5.452	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042022\
Data File : VN072107.D
Acq On : 21 Apr 2022 03:54
Operator : JC\MD
Sample : N2482-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-23R

Quant Time: Apr 21 07:14:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

