

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042022\
 Data File : VN072112.D
 Acq On : 21 Apr 2022 05:55
 Operator : JC\MD
 Sample : N2482-17
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPN-4

Quant Time: Apr 21 07:16:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	621059	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1074915	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1145691	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	428234	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	330276	46.583	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.160%
35) Dibromofluoromethane	8.022	113	304186	48.091	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.180%
50) Toluene-d8	10.439	98	1375091	53.143	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.280%
62) 4-Bromofluorobenzene	12.727	95	479370	57.283	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	114.560%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	58579	19.333	ug/l	94
25) 2-Butanone	7.340	43	7454	1.706	ug/l	94
37) Ethyl Acetate	7.416	43	64053	6.625	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042022\
Data File : VN072112.D
Acq On : 21 Apr 2022 05:55
Operator : JC\MD
Sample : N2482-17
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPN-4

Quant Time: Apr 21 07:16:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

