

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042023\
 Data File : VN077395.D
 Acq On : 20 Apr 2023 13:08
 Operator : JC\MD
 Sample : PB152199ZHE#17
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152199ZHE#17

Quant Time: Apr 21 02:39:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	554334	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1024177	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	928792	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	331717	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	334115	44.100	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.200%
35) Dibromofluoromethane	8.172	113	301151	45.932	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.860%
50) Toluene-d8	10.572	98	1241736	47.046	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.100%
62) 4-Bromofluorobenzene	12.854	95	385289	43.891	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.780%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	34811	12.462	ug/l	89
20) Methylene Chloride	5.284	84	37905	4.805	ug/l	93
37) Ethyl Acetate	7.572	43	92886	10.706	ug/l	94
43) Isopropyl Acetate	8.695	43	450011	28.109	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042023\
Data File : VN077395.D
Acq On : 20 Apr 2023 13:08
Operator : JC\MD
Sample : PB152199ZHE#17
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152199ZHE#17

Quant Time: Apr 21 02:39:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

