

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042023\
 Data File : VN077397.D
 Acq On : 20 Apr 2023 13:57
 Operator : JC\MD
 Sample : 02389-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPB-1WC-23-04-17

Quant Time: Apr 21 02:39:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	541157	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1005559	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	910804	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	338651	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	345693	46.739	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	93.480%		
35) Dibromofluoromethane	8.178	113	310641	48.257	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.520%		
50) Toluene-d8	10.572	98	1248163	48.165	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	96.320%		
62) 4-Bromofluorobenzene	12.854	95	409111	47.467	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	94.940%		
Target Compounds					Qvalue	
16) Acetone	4.443	43	51097	18.737	ug/l #	81
20) Methylene Chloride	5.290	84	40592	5.271	ug/l	96
37) Ethyl Acetate	7.572	43	81773	9.600	ug/l #	94
43) Isopropyl Acetate	8.695	43	402139	25.584	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042023\
Data File : VN077397.D
Acq On : 20 Apr 2023 13:57
Operator : JC\MD
Sample : 02389-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPB-1WC-23-04-17

Quant Time: Apr 21 02:39:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

