

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042023\
 Data File : VN077403.D
 Acq On : 20 Apr 2023 16:22
 Operator : JC\MD
 Sample : 02389-29
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPB-8WC-23-04-17

Quant Time: Apr 21 02:40:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	545128	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1022511	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	929761	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	336567	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	338866	45.483	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.960%
35) Dibromofluoromethane	8.172	113	308929	47.195	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.380%
50) Toluene-d8	10.572	98	1254416	47.604	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.200%
62) 4-Bromofluorobenzene	12.854	95	413510	47.182	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.360%
Target Compounds						
					Qvalue	
16) Acetone	4.454	43	40965	14.913	ug/l	94
20) Methylene Chloride	5.290	84	40020	5.159	ug/l	94
37) Ethyl Acetate	7.572	43	77885	8.992	ug/l	97
43) Isopropyl Acetate	8.695	43	359020	22.462	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042023\
Data File : VN077403.D
Acq On : 20 Apr 2023 16:22
Operator : JC\MD
Sample : 02389-29
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPB-8WC-23-04-17

Quant Time: Apr 21 02:40:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

