

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042023\
 Data File : VN077404.D
 Acq On : 20 Apr 2023 16:46
 Operator : JC\MD
 Sample : 02390-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPC-1WC-23-04-17

Quant Time: Apr 21 02:40:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	537827	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1023565	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	976272	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	353052	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	332027	45.170	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.340%
35) Dibromofluoromethane	8.172	113	305630	46.643	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.280%
50) Toluene-d8	10.572	98	1259821	47.760	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.520%
62) 4-Bromofluorobenzene	12.854	95	423446	48.266	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.540%
Target Compounds						
					Qvalue	
16) Acetone	4.437	43	41566	15.337	ug/l	93
20) Methylene Chloride	5.290	84	40827	5.334	ug/l	100
37) Ethyl Acetate	7.566	43	73141	8.435	ug/l #	94
43) Isopropyl Acetate	8.695	43	394032	24.627	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042023\
Data File : VN077404.D
Acq On : 20 Apr 2023 16:46
Operator : JC\MD
Sample : 02390-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPC-1WC-23-04-17

Quant Time: Apr 21 02:40:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

