

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042023\
 Data File : VN077406.D
 Acq On : 20 Apr 2023 17:35
 Operator : JC\MD
 Sample : 02390-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPC-3WC-23-04-17

Quant Time: Apr 21 02:41:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	573828	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1052120	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	974766	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	344080	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	345502	44.054	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.100%
35) Dibromofluoromethane	8.172	113	317271	47.105	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.220%
50) Toluene-d8	10.571	98	1315391	48.513	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.020%
62) 4-Bromofluorobenzene	12.854	95	422728	46.877	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.760%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	36988	12.791	ug/l	98
20) Methylene Chloride	5.295	84	35798	4.384	ug/l	93
37) Ethyl Acetate	7.572	43	75815	8.506	ug/l	97
43) Isopropyl Acetate	8.695	43	401947	24.440	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042023\
Data File : VN077406.D
Acq On : 20 Apr 2023 17:35
Operator : JC\MD
Sample : 02390-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPC-3WC-23-04-17

Quant Time: Apr 21 02:41:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

