



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042023\
Data File : VN077388.D
Acq On : 20 Apr 2023 09:43
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Title : SW846 8260
Last Update : Thu Apr 06 08:27:03 2023



AutoFind: Scans 1854, 1855, 1856; Background Corrected with Scan 1846

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	17.9	22915	PASS
75	95	30	60	54.0	69245	PASS
95	95	100	100	100.0	128149	PASS
96	95	5	9	6.6	8414	PASS
173	174	0.00	2	0.9	823	PASS
174	95	50	100	69.5	89013	PASS
175	174	5	9	6.7	5921	PASS
176	174	95	101	96.5	85907	PASS
177	176	5	9	6.6	5696	PASS

Average of 12.842 to 12.854 min.: VN077388.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
35.90	558	47.10	614	61.05	5701	71.95	977
36.10	421	47.90	805	62.00	6511	73.00	623
37.05	5748	49.00	5238	62.95	5466	74.05	2525
38.05	4805	50.00	22915	63.95	527	75.00	6924
38.90	696	51.05	6966	64.90	72	76.00	6085
39.10	1612	52.00	565	65.10	53	76.95	424
42.75	180	55.00	397	66.85	144	77.20	425
44.00	837	56.00	2243	67.10	300	77.90	478
45.05	1401	56.90	3700	68.00	14152	78.85	6920
45.85	133	58.05	137	69.00	14031	79.90	1595
46.90	838	59.95	1149	70.00	1210	80.90	6723

Average of 12.842 to 12.854 min.: VN077388.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
81.90	1422	96.05	8414	112.85	224	134.70	59
82.65	155	96.95	363	115.85	596	135.70	62
83.00	132	103.85	458	116.95	1127	136.95	283
87.05	5142	104.10	366	117.95	701	138.90	52
88.00	5355	105.05	328	118.90	997	139.70	78
90.80	205	105.90	1072	124.80	76	140.00	62
91.05	344	109.60	76	127.95	300	140.90	1464
92.00	4115	110.70	150	129.05	194	141.85	238
93.00	5565	111.00	157	129.60	240	143.00	1579
94.05	16773	111.70	144	130.00	353	145.90	239
95.00	128149	112.10	63	130.80	152	146.80	103

Average of 12.842 to 12.854 min.: VN077388.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
147.80	154	174.00	89013				
148.00	418	175.05	5921				
149.00	56	176.00	85907				
152.90	111	177.00	5696				
153.95	112	177.90	70				
154.95	235	283.30	62				
157.00	345						
159.05	233						
160.80	117						
172.20	81						
173.00	823						