

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042121\
 Data File : VN066699.D
 Acq On : 21 Apr 2021 15:20
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Apr 21 16:52:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 21 16:22:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.588	168	310515	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.514	114	483269	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	424894	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.287	152	186573	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.950	65	181763	47.91	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.82%
35) Dibromofluoromethane	7.508	113	148597	49.10	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.20%
50) Toluene-d8	10.026	98	583343	49.43	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.86%
62) 4-Bromofluorobenzene	12.346	95	181734	44.96	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.92%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042121\
Data File : VN066699.D
Acq On : 21 Apr 2021 15:20
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Apr 21 16:52:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 21 16:22:57 2021
Response via : Initial Calibration

