

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042121\
 Data File : VN066702.D
 Acq On : 21 Apr 2021 17:35
 Operator : JC/MD
 Sample : VN0421WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0421WBL01

Quant Time: Apr 22 08:29:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 22 08:25:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	308609	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	479757	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.349	117	427249	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	187042	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	183370	48.64	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.28%
35) Dibromofluoromethane	7.513	113	149913	49.90	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.80%
50) Toluene-d8	10.029	98	576223	49.18	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.36%
62) 4-Bromofluorobenzene	12.346	95	180966	45.10	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.20%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042121\
Data File : VN066702.D
Acq On : 21 Apr 2021 17:35
Operator : JC/MD
Sample : VN0421WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0421WBL01

Quant Time: Apr 22 08:29:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 22 08:25:55 2021
Response via : Initial Calibration

