


1,2,3-Trichlorobenzene

Response Ratio



$$\text{Response} = 9.110\text{e-}001 * \text{Amt} - 2.475\text{e-}002$$


Coef of Det (r^2) = 0.997880 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N042121W.M

Calibration Table Last Updated: Thu Apr 22 08:25:55 2021

2-Chloroethyl Vinyl ether

Response Ratio



$$\text{Response} = 1.024\text{e-}001 * \text{Amt} - 4.537\text{e-}002$$


Coef of Det (r^2) = 0.989828 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N042121W.M

Calibration Table Last Updated: Thu Apr 22 08:25:55 2021

Ethyl methacrylate

Response Ratio



$$\text{Response} = 5.889\text{e-}001 * \text{Amt} - 3.709\text{e-}002$$


Coef of Det (r^2) = 0.996271 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N042121W.M

Calibration Table Last Updated: Thu Apr 22 08:25:55 2021

Naphthalene


Response Ratio



Response = 2.392e+000 * Amt - 8.831e-002
 Coef of Det (r^2) = 0.996160 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N042121W.M
 Calibration Table Last Updated: Thu Apr 22 08:25:55 2021

1,2,4-Trichlorobenzene

Response Ratio




$$\text{Response} = 9.844\text{e-}001 * \text{Amt} - 3.940\text{e-}002$$

Coef of Det (r^2) = 0.996781 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N042121W.M

Calibration Table Last Updated: Thu Apr 22 08:25:55 2021


trans-1,4-Dichloro-2-butene



Response = $3.925 \times 10^{-1} \times \text{Amt} - 4.310 \times 10^{-2}$
Coef of Det (r^2) = 0.995003 Curve Fit: Linear
Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N042121W.M
Calibration Table Last Updated: Thu Apr 22 08:25:55 2021

2-Chlorotoluene


Response Ratio



$\text{Response} = 2.705\text{e}+000 * \text{Amt} - 5.063\text{e}-002$
 Coef of Det (r^2) = 0.999340 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N042121W.M
 Calibration Table Last Updated: Thu Apr 22 08:25:55 2021

Styrene


Response Ratio



Response = 1.284e+000 * Amt - 7.650e-002
Coef of Det (r^2) = 0.996984 Curve Fit: Linear
Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N042121W.M
Calibration Table Last Updated: Thu Apr 22 08:25:55 2021

2-Hexanone

Response Ratio



$$\text{Response} = 3.507\text{e-}001 * \text{Amt} - 1.167\text{e-}001$$

Coef of Det (r^2) = 0.996535 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N042121W.M

Calibration Table Last Updated: Thu Apr 22 08:25:55 2021