

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042122\
 Data File : VN072126.D
 Acq On : 21 Apr 2022 15:33
 Operator : JC\MD
 Sample : N2482-20
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB

Quant Time: Apr 22 01:24:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	441271	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	687986	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	594515	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	187837	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	262031	52.015	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.020%	
35) Dibromofluoromethane	8.022	113	225449	55.688	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 111.380%	
50) Toluene-d8	10.439	98	829455	50.085	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 100.160%	
62) 4-Bromofluorobenzene	12.727	95	228954	42.746	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 85.500%	
Target Compounds						
58) 2-Chloroethyl Vinyl ether	10.033	63	1446	17.627	ug/l	Qvalue # 70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042122\
Data File : VN072126.D
Acq On : 21 Apr 2022 15:33
Operator : JC\MD
Sample : N2482-20
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB

Quant Time: Apr 22 01:24:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

