

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN042122\  
 Data File : VN072127.D  
 Acq On : 21 Apr 2022 15:57  
 Operator : JC\MD  
 Sample : N2482-19  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 EB-2022-04-19

Quant Time: Apr 22 01:24:24 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N033022W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Mar 31 01:57:54 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Pentafluorobenzene       | 8.081  | 168   | 477690   | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.963  | 114   | 764257   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 11.739 | 117   | 622062   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.669 | 152   | 203618   | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 8.434  | 65    | 262678   | 48.168   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 61 - 141 | Recovery | =      | 96.340%  |
| 35) Dibromofluoromethane    | 8.016  | 113   | 227288   | 50.540   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 69 - 133 | Recovery | =      | 101.080% |
| 50) Toluene-d8              | 10.439 | 98    | 787143   | 42.786   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 65 - 126 | Recovery | =      | 85.580%  |
| 62) 4-Bromofluorobenzene    | 12.727 | 95    | 227671   | 38.265   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 58 - 135 | Recovery | =      | 76.520%  |
| Target Compounds            |        |       |          |          |        |          |
|                             |        |       |          |          | Qvalue |          |
| 16) Acetone                 | 4.287  | 43    | 43893    | 18.833   | ug/l   | 95       |
| 25) 2-Butanone              | 7.345  | 43    | 11691    | 3.480    | ug/l   | 97       |
| 29) Tetrahydrofuran         | 7.692  | 42    | 5217     | 2.398    | ug/l   | 92       |
| -----                       |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN042122\  
Data File : VN072127.D  
Acq On : 21 Apr 2022 15:57  
Operator : JC\MD  
Sample : N2482-19  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
EB-2022-04-19

Quant Time: Apr 22 01:24:24 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N033022W.M  
Quant Title : SW846 8260  
QLast Update : Thu Mar 31 01:57:54 2022  
Response via : Initial Calibration

